

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q1447
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1447 **SAS No.:** Q1447

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166897BS							
Arsenic	ug/L	4000	3960		99	80 - 120	P
Barium	ug/L	1000	877		88	80 - 120	P
Cadmium	ug/L	1000	931		93	80 - 120	P
Chromium	ug/L	2000	2010		100	80 - 120	P
Lead	ug/L	5000	4730		95	80 - 120	P
Selenium	ug/L	10000	9740		97	80 - 120	P
Silver	ug/L	380	377		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q1447

Contract: GENV01

Lab Code: CHEM

Case No.: Q1447

SAS No.: Q1447

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166936BS Mercury	ug/L	4.0	4.09		102	80 - 120	CV