

Continuing Calibration Report for SequenceID : FC030425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068360.D

Aliphatic C9-C12	8895633.000	60.000	3.101	6.366	148260.550	141035.502	-5.123
Aliphatic C12-C16	5648971.000	40.000	6.367	9.750	141224.275	136654.515	-3.344
Aliphatic C16-C21	8146363.000	60.000	9.751	13.103	135772.717	131630.895	-3.147
Aliphatic C21-C28	10085167.000	80.000	13.104	16.756	126064.588	122691.322	-2.749
Aliphatic C28-C40	12598792.000	120.000	16.757	21.558	104989.933	102846.462	-2.084
Aliphatic EPH	45374926.000	360.000	3.101	21.558	126041.461	122175.127	-3.165