

Continuing Calibration Report for SequenceID : FC030425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068365.D

Aliphatic C9-C12	9010317.000	60.000	3.101	6.366	150171.950	141035.502	-6.478
Aliphatic C12-C16	5674412.000	40.000	6.367	9.750	141860.300	136654.515	-3.809
Aliphatic C16-C21	8142808.000	60.000	9.751	13.103	135713.467	131630.895	-3.102
Aliphatic C21-C28	10124899.000	80.000	13.104	16.756	126561.238	122691.322	-3.154
Aliphatic C28-C40	12735293.000	120.000	16.757	21.558	106127.442	102846.462	-3.190
Aliphatic EPH	45687729.000	360.000	3.101	21.558	126910.358	122175.127	-3.876