

Continuing Calibration Report for SequenceID : FD030325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049152.D

Aromatic C10-C12	7166334.000	40.000	4.074	5.792	179158.350	159020.606	-12.664
Aromatic C12-C16	10830604.000	60.000	5.793	8.394	180510.067	160935.866	-12.163
Aromatic C16-C21	13356833.000	80.000	8.395	12.657	166960.413	153359.309	-8.869
Aromatic C21-C36	26081586.000	180.000	12.658	18.067	144897.700	143114.753	-1.246
Aromatic EPH	57435357.000	360.000	4.074	18.067	159542.658	150128.823	-6.271