

Continuing Calibration Report for SequenceID : FD030325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049155.D

Aromatic C10-C12	7286344.000	40.000	4.074	5.792	182158.600	159020.606	-14.550
Aromatic C12-C16	11027649.000	60.000	5.793	8.394	183794.150	160935.866	-14.203
Aromatic C16-C21	13539680.000	80.000	8.395	12.657	169246.000	153359.309	-10.359
Aromatic C21-C36	27644418.000	180.000	12.658	18.067	153580.100	143114.753	-7.313
Aromatic EPH	59498091.000	360.000	4.074	18.067	165272.475	150128.823	-10.087