

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** G Environmental
Lab Code: CHEM **Cas No:** Q1447 **SAS No :** Q1447 **SDG No:** Q1447
Sample No : PB166923BS **Datafile:** FC068323.D
Client ID : PB166923BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	6.53	65		(40-140)
Aliphatic C12-C16	6.7	5.75	86		(40-140)
Aliphatic C16-C21	10	9.57	96		(40-140)
Aliphatic C21-C28	13.3	12.9	97		(40-140)
Aliphatic C28-C40	20.0	21.3	106		(40-140)