

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf030425

SequenceID :	<u>bf030425</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)		_____	_____	<u>✓</u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)		_____	_____	<u>✓</u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series		_____	_____	<u>✓</u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series		_____	_____	<u>✓</u>
5. GC/MS Calibration Met:		_____	_____	<u>✓</u>
a. Initial calibration Meet Criteria		_____	_____	<u>✓</u>
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____
b. Continuous Calibration(CCC) Meet Criteria		_____	<u>✓</u>	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____
Compounds #56,62 & 77 are biased high in SSTDCCC (BF141824.D). If any samples are found with Hit of this compound they will be analyzed again with proper passing CCC.		_____	_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		_____	<u>✓</u>	_____
a. B/N Fraction		_____	_____	_____
d. Acid Fraction		_____	_____	_____

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

✓

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

✓

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

✓

ADDITIONAL COMMENTS:

The Recovery is biased high for compound #62 in PB166957BS. This data will be used for hardcopies.

The sample Q1480-01 had to be analyzed with 2X dilution and the sample Q1480-09 had to be analyzed with 5X dilution due to dirty and viscous matrix.

krishna

Analyst

03/05/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf030425

SequenceID : bf030425 NA NO YES