

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01  
 Lab Code: CHEM Case No.: Q1447 SAS No.: Q1447 SDG NO.: Q1447  
 Lab File ID: VY021406.D Date Analyzed: 03/05/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 11:39  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 179824        | 13.346 |  |  |  |  |
| UPPER LIMIT    | 359648        | 13.846 |  |  |  |  |
| LOWER LIMIT    | 89912         | 12.846 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| WC1            | 164565        | 13.35  |  |  |  |  |
| VY0305SBL01    | 140807        | 13.35  |  |  |  |  |
| VY0305SBS01    | 150751        | 13.35  |  |  |  |  |
| VY0305SBSD01   | 157165        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.