

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1447 SAS No.: Q1447 SDG NO.: Q1447
Lab File ID: VY021395.D BFB Injection Date: 03/04/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:52
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.6
75	30.0 - 60.0% of mass 95	54.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1.4 (1.8) 1
174	50.0 - 100.0% of mass 95	77.8
175	5.0 - 9.0% of mass 174	6.6 (8.4) 1
176	95.0 - 101.0% of mass 174	75.7 (97.2) 1
177	5.0 - 9.0% of mass 176	5 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC010	VSTDIC010	VY021397.D	03/04/2025	09:46
VSTDIC020	VSTDIC020	VY021398.D	03/04/2025	10:09
VSTDIC050	VSTDIC050	VY021399.D	03/04/2025	10:30
VSTDIC100	VSTDIC100	VY021400.D	03/04/2025	11:06
VSTDIC150	VSTDIC150	VY021401.D	03/04/2025	11:29
VSTDIC005	VSTDIC005	VY021403.D	03/04/2025	12:15