

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Mahesh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133187,VP133189
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199
CCC	VP133188,VP133190
Internal Standard/PEM	
ICV/I.BLK	VP133200
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045067.D	28 Feb 2025 01:03		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045068.D	28 Feb 2025 01:27		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045069.D	28 Feb 2025 02:13		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045070.D	28 Feb 2025 02:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045071.D	28 Feb 2025 03:00		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045072.D	28 Feb 2025 03:23		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045073.D	28 Feb 2025 03:47		JC/MD	Ok,M
8	IBLK	IBLK	VX045074.D	28 Feb 2025 04:10		JC/MD	Ok
9	VSTDICV050	ICVVX022825	VX045075.D	28 Feb 2025 04:33		JC/MD	Ok,M
10	BFB	BFB	VX045076.D	28 Feb 2025 10:03		JC/MD	Ok
11	VSTDCCC050	VSTDCCC050	VX045077.D	28 Feb 2025 10:32	V12668	JC/MD	Ok,M
12	VX0228MBL01	VX0228MBL01	VX045078.D	28 Feb 2025 11:00		JC/MD	Ok
13	VX0228WBL01	VX0228WBL01	VX045079.D	28 Feb 2025 11:23		JC/MD	Ok
14	VX0228WBS01	VX0228WBS01	VX045080.D	28 Feb 2025 11:46		JC/MD	Ok,M
15	VX0228WBSD01	VX0228WBSD01	VX045081.D	28 Feb 2025 12:13		JC/MD	Ok,M
16	Q1401-03	BP-VPB-192-GW-840-8	VX045082.D	28 Feb 2025 12:37	vial B pH<2	JC/MD	Ok
17	Q1401-06	BP-VPB-192-GW-900-9	VX045083.D	28 Feb 2025 13:00	vial B pH<2	JC/MD	Ok
18	Q1423-01	BP-VPB-192-TB-20250	VX045084.D	28 Feb 2025 13:23	vial B pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Maresh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133187,VP133189		
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199		
CCC	VP133188,VP133190		
Internal Standard/PEM			
ICV/I.BLK	VP133200		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1423-03	BP-VPB-192-EB-20250	VX045085.D	28 Feb 2025 13:47	vial B pH<2 EB	JC/MD	Ok
20	Q1403-01	Storage-Blank-SOIL-RE	VX045086.D	28 Feb 2025 14:10	vial B pH<2	JC/MD	Ok
21	Q1403-02	Storage-Blank-WATER-	VX045087.D	28 Feb 2025 14:33	vial B pH<2	JC/MD	Ok
22	Q1435-01	286107	VX045088.D	28 Feb 2025 14:57	vial A pH<2	JC/MD	Ok
23	Q1403-04	Storage-Blank-SAMPLE	VX045089.D	28 Feb 2025 15:20	vial B pH<2	JC/MD	Ok
24	Q1462-02	FB	VX045090.D	28 Feb 2025 15:44	vial A pH<2 FB	JC/MD	Ok
25	Q1469-01	Storage-Blank-SOIL-RE	VX045091.D	28 Feb 2025 16:07	vial A pH<2	JC/MD	Ok
26	Q1469-02	Storage-Blank-WATER-	VX045092.D	28 Feb 2025 16:31	vial A pH<2	JC/MD	Ok
27	Q1469-03	Storage-Blank-WATER-	VX045093.D	28 Feb 2025 16:54	vial A pH<2	JC/MD	Ok
28	Q1469-04	Storage-Blank-SAMPLE	VX045094.D	28 Feb 2025 17:17	vial A pH<2	JC/MD	Ok
29	Q1403-03	Storage-Blank-WATER-	VX045095.D	28 Feb 2025 17:41	vial B pH<2	JC/MD	Ok
30	Q1462-01	MW2	VX045096.D	28 Feb 2025 18:04	vial A pH<2	JC/MD	Ok,M
31	IBLK	IBLK	VX045097.D	28 Feb 2025 18:27		JC/MD	Ok
32	VSTDCCC050	VSTDCCC050EC	VX045098.D	28 Feb 2025 18:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030525

Review By	John Carlone	Review On	3/6/2025 8:28:41 AM
Supervise By	Maresh Dadoda	Supervise On	3/6/2025 2:56:36 PM
SubDirectory	VX030525	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133217
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133218,VP133219

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045126.D	05 Mar 2025 08:40		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045127.D	05 Mar 2025 09:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0305MBL01	VX0305MBL01	VX045128.D	05 Mar 2025 09:39		JC/MD	Ok
4	VX0305WBL01	VX0305WBL01	VX045129.D	05 Mar 2025 10:03		JC/MD	Ok
5	VX0305WBS01	VX0305WBS01	VX045130.D	05 Mar 2025 10:26		JC/MD	Ok,M
6	VX0305MBS01	VX0305MBS01	VX045131.D	05 Mar 2025 10:53		JC/MD	Ok,M
7	Q1480-17	TP-3	VX045132.D	05 Mar 2025 11:16		JC/MD	Ok
8	Q1448-04ME	P3ME	VX045133.D	05 Mar 2025 11:40		JC/MD	Ok
9	IBLK	IBLK	VX045134.D	05 Mar 2025 12:03		JC/MD	Ok
10	VX0305WBSD01	VX0305WBSD01	VX045135.D	05 Mar 2025 12:26		JC/MD	Ok,M
11	Q1478-04DL	IDW-AQ-DRUM-616-02	VX045136.D	05 Mar 2025 12:50	vial B pH<2	JC/MD	Ok
12	Q1478-06DL	IDW-AQ-DRUM-614-02	VX045137.D	05 Mar 2025 13:13	vial B pH<2	JC/MD	Ok
13	Q1478-02DL	IDW-AQ-DRUM-610-02	VX045138.D	05 Mar 2025 13:36	Not req.	JC/MD	Not Ok
14	Q1478-08	IDW-AQ-DRUM-612-02	VX045139.D	05 Mar 2025 13:59	Need lower dilution	JC/MD	Not Ok
15	VX0305MBSD01	VX0305MBSD01	VX045140.D	05 Mar 2025 14:23		JC/MD	Ok,M
16	Q1448-04MEDL	P3MEDL	VX045141.D	05 Mar 2025 14:46	Not req.	JC/MD	Not Ok
17	IBLK	IBLK	VX045142.D	05 Mar 2025 15:09		JC/MD	Ok
18	Q1478-08	IDW-AQ-DRUM-612-02	VX045143.D	05 Mar 2025 15:32	Need lower dilution	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030525

Review By	John Carlone	Review On	3/6/2025 8:28:41 AM
Supervise By	Maresh Dadoda	Supervise On	3/6/2025 2:56:36 PM
SubDirectory	VX030525	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133217		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133218,VP133219		

19	Q1478-02DL	IDW-AQ-DRUM-610-02	VX045144.D	05 Mar 2025 15:56	vial B pH<2	JC/MD	Ok
20	Q1478-08	IDW-AQ-DRUM-612-02	VX045145.D	05 Mar 2025 16:19	vial A pH<2	JC/MD	Ok
21	Q1501-03	EB01-030525	VX045146.D	05 Mar 2025 16:42	vial A pH<2 EB	JC/MD	Ok
22	Q1501-07	TB01-030525	VX045147.D	05 Mar 2025 17:05	vial A pH<2 TB	JC/MD	Ok
23	Q1501-04	BR-05-465-030525	VX045148.D	05 Mar 2025 17:28	Need Straight Run	JC/MD	Not Ok
24	Q1488-15	TB03042025	VX045149.D	05 Mar 2025 17:52	vial A pH<2 TB	JC/MD	Ok
25	Q1488-13	ENV-102-GW01	VX045150.D	05 Mar 2025 18:15	vial A pH<2	JC/MD	Ok
26	Q1488-14	ENV-104-GW01	VX045151.D	05 Mar 2025 18:38	vial A pH<2 Need 4X	JC/MD	Dilution
27	Q1494-01	PURGE-WATER	VX045152.D	05 Mar 2025 19:01	vial A pH<2 E flag in previous sample	JC/MD	ReRun
28	VSTDCCC050	VSTDCCC050EC	VX045153.D	05 Mar 2025 19:24		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022525

Review By	Mahesh Dadoda	Review On	2/26/2025 1:45:35 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/26/2025 1:46:55 PM		
SubDirectory	VY022525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y022525s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133154
Initial Calibration Stds	VP133137,VP133140,VP133143,VP133146,VP133147,VP133149
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133150
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021308.D	25 Feb 2025 12:10		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021309.D	25 Feb 2025 12:40		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021310.D	25 Feb 2025 13:03		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021311.D	25 Feb 2025 13:26	Acrolein fail for method	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021312.D	25 Feb 2025 14:48		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021313.D	25 Feb 2025 15:34	Not used	SY/MD	Not Ok
7	VSTDIC150	VSTDIC150	VY021314.D	25 Feb 2025 15:57		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021315.D	25 Feb 2025 16:24		SY/MD	Ok
9	VSTDIC100	VSTDIC100	VY021316.D	25 Feb 2025 16:49		SY/MD	Ok,M
10	VSTDICV050	ICVVY022525	VY021317.D	25 Feb 2025 17:31	ICV Failed for comp. #10	SY/MD	Ok,M
11	VIBLK	VIBLK	VY021318.D	25 Feb 2025 17:54		SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Mahesh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133167 VP133165,VP133166

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021345.D	27 Feb 2025 09:41		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021346.D	27 Feb 2025 10:13		SY/MD	Ok,M
3	VY0227SBL01	VY0227SBL01	VY021347.D	27 Feb 2025 10:42		SY/MD	Ok
4	VY0227SBS01	VY0227SBS01	VY021348.D	27 Feb 2025 11:18		SY/MD	Ok,M
5	VY0227SBSD01	VY0227SBSD01	VY021349.D	27 Feb 2025 11:40		SY/MD	Ok,M
6	Q1428-01	NB-07-022525	VY021350.D	27 Feb 2025 12:04	Vial-A ISTD Fail	SY/MD	ReRun
7	Q1442-03	351	VY021351.D	27 Feb 2025 12:28	Vial-A	SY/MD	Ok
8	Q1416-01RE	OK-01-022425RE	VY021352.D	27 Feb 2025 12:51	Vial-B ISTD Fail	SY/MD	Confirms
9	Q1428-01RE	NB-07-022525RE	VY021353.D	27 Feb 2025 13:14	Vial-B Internal Standard Fail	SY/MD	Confirms
10	Q1418-01	TR-06-02242025	VY021354.D	27 Feb 2025 13:38	Vial-B Not purge	SY/MD	Not Ok
11	Q1415-03RE	B-163-SB02RE	VY021355.D	27 Feb 2025 14:01	Vial-B Internal standard fail;Not match for com.#20	SY/MD	Confirms
12	Q1422-02	B-154-SB02	VY021356.D	27 Feb 2025 14:25	Vial-A Internal standard fail	SY/MD	ReRun
13	Q1422-01	B-154-SB01	VY021357.D	27 Feb 2025 14:48	Vial-A Internal standard fail	SY/MD	ReRun
14	Q1421-09	P001-CLAY-CF02-01	VY021358.D	27 Feb 2025 15:12	Internal Standard Fail	SY/MD	ReRun
15	Q1421-07	P001-CLAY-CF01-02	VY021359.D	27 Feb 2025 15:35	Vial-A Internal standard fail	SY/MD	ReRun
16	Q1421-01	P001-CLAY-CF01-01	VY021360.D	27 Feb 2025 15:58	Vial-A Internal standard fail	SY/MD	Ok
17	Q1427-01RE	VNJ-227RE	VY021361.D	27 Feb 2025 16:22	Vial-B Internal standard fail;Surrogate fail	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Maresh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP133167		
Initial Calibration Stds	VP133165,VP133166		
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1420-07	TP-2-WC-VOC	VY021362.D	27 Feb 2025 16:45	Vial-B	SY/MD	Ok
19	Q1411-03	HR-03-02212025	VY021363.D	27 Feb 2025 17:09	Vial-B Not purge	SY/MD	Not Ok
20	Q1411-01	HR-02-02212025	VY021364.D	27 Feb 2025 17:32	Vial-B Internal standard fail	SY/MD	Ok
21	Q1440-01	OR-02-022625	VY021365.D	27 Feb 2025 17:55	Vial-A Internal standard fail	SY/MD	ReRun
22	Q1448-01	PSP1	VY021366.D	27 Feb 2025 18:19	Vial-A Internal standard fail	SY/MD	ReRun
23	Q1448-02	P1	VY021367.D	27 Feb 2025 18:42	Vial-A Internal standard fail	SY/MD	ReRun
24	Q1448-03	P2	VY021368.D	27 Feb 2025 19:06	Vial-A Internal standard fail	SY/MD	ReRun
25	Q1448-04	P3	VY021369.D	27 Feb 2025 19:29	Vial-A Internal standard fail;Need MeOH	SY/MD	Dilution
26	Q1448-05	P4	VY021370.D	27 Feb 2025 19:52	Vial-A Internal standard fail	SY/MD	ReRun
27	Q1421-02MS	P001-CLAY-CF01-01M	VY021371.D	27 Feb 2025 20:15	Vial-A Internal standard fail	SY/MD	Ok,M
28	Q1421-03MSD	P001-CLAY-CF01-01M	VY021372.D	27 Feb 2025 20:38	Vial-A Internal standard fail	SY/MD	Ok,M
29	VSTDCCC050	VSTDCCC050EC	VY021373.D	27 Feb 2025 21:01		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022825

Review By	John Carlone	Review On	3/3/2025 12:12:17 PM
Supervise By	Maresh Dadoda	Supervise On	3/3/2025 1:51:10 PM
SubDirectory	VY022825	HP Acquire Method	MSVOA_Y HP Processing Method 82y022525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133182
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133183,VP133184 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021374.D	28 Feb 2025 09:03		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021375.D	28 Feb 2025 09:44	CCAL failed low for comp. #11; CCAL failed high for comp. #16	SY/MD	Ok,M
3	VY0228SBL01	VY0228SBL01	VY021376.D	28 Feb 2025 10:59		SY/MD	Ok
4	VY0228SBS01	VY0228SBS01	VY021377.D	28 Feb 2025 11:32		SY/MD	Ok,M
5	VY0228SBSD01	VY0228SBSD01	VY021378.D	28 Feb 2025 11:55		SY/MD	Ok,M
6	Q1421-07RE	P001-CLAY-CF01-02RE	VY021379.D	28 Feb 2025 12:56	Vial-B Internal Standard Fail	SY/MD	Confirms
7	Q1422-01RE	B-154-SB01RE	VY021380.D	28 Feb 2025 13:20	Vial-B Internal Standard Fail	SY/MD	Confirms
8	Q1422-02	B-154-SB02	VY021381.D	28 Feb 2025 15:31	Vial-B	SY/MD	Ok
9	Q1448-01RE	PSP1RE	VY021382.D	28 Feb 2025 15:55	Vial-B CCAL failed low for comp. #11	SY/MD	Confirms
10	Q1448-02RE	P1RE	VY021383.D	28 Feb 2025 16:18	Vial-B CCAL failed low for comp. #11	SY/MD	Confirms
11	Q1448-03RE	P2RE	VY021384.D	28 Feb 2025 16:41	Vial-B CCAL failed low for comp. #11;Not match for comp. #20	SY/MD	Confirms
12	Q1448-05RE	P4RE	VY021385.D	28 Feb 2025 17:05	Vial-B CCAL failed low for comp. #11; CCAL failed high for comp. #16;Not match for comp. #11,16,20	SY/MD	Confirms
13	Q1440-01	OR-02-022625	VY021386.D	28 Feb 2025 17:28	Vial-B	SY/MD	Ok
14	Q1441-01	HD-02-022625	VY021387.D	28 Feb 2025 17:52	Vial-B Internal Standard Fail	SY/MD	ReRun
15	Q1458-02	SP-SOIL-VOC	VY021388.D	28 Feb 2025 18:15	vial-A	SY/MD	Ok,M

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022825

Review By	John Carlone	Review On	3/3/2025 12:12:17 PM
Supervise By	Maresh Dadoda	Supervise On	3/3/2025 1:51:10 PM
SubDirectory	VY022825	HP Acquire Method	MSVOA_Y HP Processing Method 82y022525s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133182		
CCC	VP133183,VP133184		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

16	Q1463-01	T1	VY021389.D	28 Feb 2025 18:38	Vial-A CCAL failed low for comp. #11	SY/MD	ReRun
17	Q1457-01	HR-0-02272025	VY021390.D	28 Feb 2025 19:02	Vial-A	SY/MD	Ok
18	Q1463-03	T3	VY021391.D	28 Feb 2025 19:25	Vial-A Internal Standard Fail; CCAL failed low for comp. #11; CCAL failed high for comp.#16	SY/MD	ReRun
19	Q1463-04	T4	VY021392.D	28 Feb 2025 19:49	Vial-A CCAL failed low for comp. #11	SY/MD	ReRun
20	Q1463-05	T5	VY021393.D	28 Feb 2025 20:12	Vial-A Internal Standard Fail; CCAL failed low for comp. #11; CCAL failed high for comp.#16	SY/MD	ReRun
21	VSTDCCC050	VSTDCCC050EC	VY021394.D	28 Feb 2025 20:35		SY/MD	Ok,M

M : Manual Integration