

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1448 SAS No.: Q1448 SDG NO.: Q1448
 Lab File ID: VX045127.D Date Analyzed: 03/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	112291	5.54	190470	6.75	164207	10.05
UPPER LIMIT	224582	6.044	380940	7.251	328414	10.549
LOWER LIMIT	56145.5	5.044	95235	6.251	82103.5	9.549
EPA SAMPLE NO.						
P3ME	67331	5.54	128533	6.76	115484	10.06
VX0305MBL01	68268	5.54	138238	6.76	127715	10.05
VX0305MBS01	93506	5.54	165212	6.76	145087	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.