

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1448 SAS No.: Q1448 SDG NO.: Q1448
 Lab File ID: VY021375.D Date Analyzed: 02/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	113605	13.353				
UPPER LIMIT	227210	13.853				
LOWER LIMIT	56802.5	12.853				
EPA SAMPLE NO.						
PSP1RE	76917	13.35				
P1RE	75987	13.35				
P2RE	61708	13.35				
P4RE	63405	13.35				
VY0228SBL01	59388	13.35				
VY0228SBS01	115456	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.