

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1448 SAS No.: Q1448 SDG NO.: Q1448
Lab File ID: VY021374.D BFB Injection Date: 02/28/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.7
75	30.0 - 60.0% of mass 95	56.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	1.4 (1.6) 1
174	50.0 - 100.0% of mass 95	84
175	5.0 - 9.0% of mass 174	6.5 (7.7) 1
176	95.0 - 101.0% of mass 174	80.5 (95.9) 1
177	5.0 - 9.0% of mass 176	5.2 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021375.D	02/28/2025	09:44
VY0228SBL01	VY0228SBL01	VY021376.D	02/28/2025	10:59
VY0228SBS01	VY0228SBS01	VY021377.D	02/28/2025	11:32
PSP1RE	Q1448-01RE	VY021382.D	02/28/2025	15:55
P1RE	Q1448-02RE	VY021383.D	02/28/2025	16:18
P2RE	Q1448-03RE	VY021384.D	02/28/2025	16:41
P4RE	Q1448-05RE	VY021385.D	02/28/2025	17:05