

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN021825

Review By	John Carlone	Review On	2/19/2025 9:44:55 AM
Supervise By	Maresh Dadoda	Supervise On	2/19/2025 1:34:51 PM
SubDirectory	VN021825	HP Acquire Method	HP Processing Method 82N021825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133068 VP133070,VP133071,VP133072,VP133073,VP133074,VP133075 VP133076

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085771.D	18 Feb 2025 10:35		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN085772.D	18 Feb 2025 11:09	Comp.#43 is on Linear Regression	JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN085773.D	18 Feb 2025 11:32	Comp.#56,58,70,86 is on Quadratic Regression	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN085774.D	18 Feb 2025 11:56	Not used	JC\MD	Not Ok
5	VSTDICC010	VSTDICC010	VN085775.D	18 Feb 2025 12:20		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN085776.D	18 Feb 2025 12:43	%D failed for Comp. #58 in 01PPB, 05PPB and 20PPB	JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN085777.D	18 Feb 2025 13:07		JC\MD	Ok,M
8	IBLK	IBLK	VN085778.D	18 Feb 2025 13:54		JC\MD	Ok
9	VSTDICC020	VSTDICC020	VN085779.D	18 Feb 2025 14:18		JC\MD	Ok,M
10	VSTDICV050	ICVVN021825	VN085780.D	18 Feb 2025 16:28	ICV Failed for comp. #95	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN022725

Review By	John Carlone	Review On	2/28/2025 10:02:38 AM
Supervise By	Mahesh Dadoda	Supervise On	2/28/2025 11:09:18 AM
SubDirectory	VN022725	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133168		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133169,VP133170		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085879.D	27 Feb 2025 09:45		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085880.D	27 Feb 2025 10:18	V13516	JC\MD	Ok,M
3	VN0227MBL01	VN0227MBL01	VN085881.D	27 Feb 2025 10:53		JC\MD	Ok
4	VN0227WBL01	VN0227WBL01	VN085882.D	27 Feb 2025 11:17		JC\MD	Ok
5	VN0227WBS01	VN0227WBS01	VN085883.D	27 Feb 2025 12:17		JC\MD	Ok,M
6	VN0227WBSD01	VN0227WBSD01	VN085884.D	27 Feb 2025 12:51	BSD failed high for comp. #11,25,52,68,69	JC\MD	Ok,M
7	Q1449-01	MW2	VN085885.D	27 Feb 2025 13:15	vial B pH<2	JC\MD	Ok,M
8	IBLK	IBLK	VN085886.D	27 Feb 2025 13:39		JC\MD	Ok
9	PB166880TB	PB166880TB	VN085887.D	27 Feb 2025 14:04		JC\MD	Ok
10	PB166880ZHE#01	PB166880ZHE#01	VN085888.D	27 Feb 2025 14:28		JC\MD	Ok
11	PB166880ZHE#02	PB166880ZHE#02	VN085889.D	27 Feb 2025 14:53		JC\MD	Ok
12	PB166880ZHE#03	PB166880ZHE#03	VN085890.D	27 Feb 2025 15:17		JC\MD	Ok
13	Q1420-04	TP-1-WC	VN085891.D	27 Feb 2025 15:41	vial A pH#5.0	JC\MD	Ok,M
14	Q1420-08	TP-2-WC	VN085892.D	27 Feb 2025 16:05	vial A pH#5.0	JC\MD	Ok
15	Q1427-02	VNJ-227	VN085893.D	27 Feb 2025 16:30	vial A pH#5.0	JC\MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VN085894.D	27 Feb 2025 16:54		JC\MD	Ok,M

M : Manual Integration