

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1449 SAS No.: Q1449 SDG NO.: Q1449
Lab File ID: VN085880.D Date Analyzed: 02/27/2025
Instrument ID: MSVOA_N Time Analyzed: 10:18
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	237819	8.22	375919	9.10	353534	11.87
UPPER LIMIT	475638	8.724	751838	9.6	707068	12.365
LOWER LIMIT	118910	7.724	187960	8.6	176767	11.365
EPA SAMPLE NO.						
MW2	272528	8.22	464123	9.10	417159	11.87
VN0227WBL01	204202	8.22	390858	9.10	343319	11.87
VN0227WBS01	262348	8.22	421598	9.10	376829	11.87
VN0227WBSD01	230188	8.22	375362	9.10	334047	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.