

Matrix Spike Summary

Client:	Ardmore Chemical	SDG No.:	Q1456
Project:	PVSC Monthly 2025	Sample ID:	Q1456-01
Client ID:	EFF-WASTE WATERMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.036		0.00093	U	0.04	1	90		02/28/2025

Matrix Spike Summary

Client:	Ardmore Chemical	SDG No.:	Q1456
Project:	PVSC Monthly 2025	Sample ID:	Q1456-01
Client ID:	EFF-WASTE WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.036		0.00093	U	0.04	1	90		02/28/2025