



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ardmore Chemical SDG No.: Q1456
Contract: ARDM01 Lab Code: CHEM Case No.: Q1456 SAS No.: Q1456
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.64	6.0	94	40 - 160	P	03/03/2025	19:30	LB134881
	Copper	21.8	20.0	109	40 - 160	P	03/03/2025	19:30	LB134881
	Lead	11.5	12.0	96	40 - 160	P	03/03/2025	19:30	LB134881
	Nickel	37.4	40.0	94	40 - 160	P	03/03/2025	19:30	LB134881
	Zinc	40.7	40.0	102	40 - 160	P	03/03/2025	19:30	LB134881
CRI01	Cadmium	5.97	6.0	100	40 - 160	P	03/04/2025	10:22	LB134892
	Copper	22.7	20.0	114	40 - 160	P	03/04/2025	10:22	LB134892
	Lead	11.7	12.0	98	40 - 160	P	03/04/2025	10:22	LB134892
	Nickel	39.7	40.0	99	40 - 160	P	03/04/2025	10:22	LB134892
	Zinc	42.5	40.0	106	40 - 160	P	03/04/2025	10:22	LB134892
CRA	Mercury	0.053	0.05	106	40 - 160	CV	03/04/2025	15:52	LB134890
CRI01	Cadmium	5.67	6.0	94	40 - 160	P	03/10/2025	12:01	LB134965
	Copper	21.4	20.0	107	40 - 160	P	03/10/2025	12:01	LB134965
	Lead	12.0	12.0	100	40 - 160	P	03/10/2025	12:01	LB134965
	Nickel	37.9	40.0	95	40 - 160	P	03/10/2025	12:01	LB134965
	Zinc	41.3	40.0	103	40 - 160	P	03/10/2025	12:01	LB134965