

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical **SDG No.:** Q1456
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q1456 **SAS No.:** Q1456

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166926BS							
Cadmium	ug/L	100	97.2		97	85 - 115	P
Copper	ug/L	150	160		107	85 - 115	P
Lead	ug/L	500	490		98	85 - 115	P
Nickel	ug/L	250	253		101	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical **SDG No.:** Q1456
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q1456 **SAS No.:** Q1456

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166974BS Mercury	ug/L	4.0	3.85		96	85 - 115	CV