

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1456 SAS No.: Q1456 SDG NO.: Q1456

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/28/2025

Lab File ID: BM049675.D Time Analyzed: 10:38

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	259599	7.804	912828	10.60	573813	14.45
UPPER LIMIT	519198	8.304	1825660	11.104	1147630	14.945
LOWER LIMIT	129800	7.304	456414	10.104	286907	13.945
EPA SAMPLE NO.						
01 PB166933BS	268342	7.81	912905	10.60	538676	14.45
02 PB166933BL	267882	7.81	881999	10.60	497065	14.44
03 PB166933BSD	281345	7.80	926819	10.60	548450	14.45
04 EFF-WASTE WATER	262882	7.80	956753	10.60	659772	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.