

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ARDM01

Lab Code: CHEM Case No.: Q1456 SAS No.: Q1456 SDG No.: Q1456

Instrument ID: MSVOA_X Calibration Date/Time: 03/04/2025 08:46

Lab File ID: VX045119.D Init. Calib. Date(s): 02/21/2025 02/21/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:58 11:29

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Chloromethane	3.029	2.998		-1.02	
Vinyl Chloride	2.923	2.622	0.1	-10.3	
Bromomethane	1.017	1.108	0.1	8.95	
Chloroethane	1.583	1.518		-4.11	
Trichlorofluoromethane	3.928	3.772		-3.97	
1,1-Dichloroethene	2.279	2.242	0.1	-1.62	
Acrolein	0.530	0.559		5.47	
Acrylonitrile	1.293	1.281		-0.93	
Methylene Chloride	2.563	2.540		-0.9	
trans-1,2-Dichloroethene	2.306	2.177		-5.59	
1,1-Dichloroethane	4.555	4.331	0.2	-4.92	
Carbon Tetrachloride	0.564	0.606	0.1	7.45	
Chloroform	4.400	4.284	0.2	-2.64	
1,1,1-Trichloroethane	0.669	0.702	0.1	4.93	
Benzene	1.731	1.805	0.5	4.28	
1,2-Dichloroethane	0.592	0.648	0.1	9.46	
Trichloroethene	0.405	0.420	0.3	3.7	
1,2-Dichloropropane	0.430	0.446		3.72	
Bromodichloromethane	0.614	0.656	0.2	6.84	
Toluene	1.987	2.055	0.4	3.42	
t-1,3-Dichloropropene	0.591	0.568	0.1	-3.89	
cis-1,3-Dichloropropene	0.670	0.675	0.2	0.75	
1,1,2-Trichloroethane	0.403	0.436	0.1	8.19	
2-Chloroethyl vinyl ether	0.325	0.331		1.85	
Dibromochloromethane	0.445	0.474	0.1	6.52	
Tetrachloroethene	0.369	0.393	0.2	6.5	
Chlorobenzene	1.231	1.270	0.5	3.17	
Ethyl Benzene	2.176	2.230	0.1	2.48	
m/p-Xylenes	0.811	0.845	0.3	4.19	
o-Xylene	0.795	0.829	0.3	4.28	
Bromoform	0.279	0.295	0.1	5.74	
1,1,2,2-Tetrachloroethane	0.675	0.729	0.3	8	
1,3-Dichlorobenzene	0.879	0.934	0.2	6.26	
1,4-Dichlorobenzene	0.867	0.909	0.2	4.84	
1,2-Dichlorobenzene	0.863	0.934	0.2	8.23	
1,2-Dichloroethane-d4	2.914	2.808	0.01	-3.64	
Toluene-d8	1.660	1.641	0.01	-1.14	
4-Bromofluorobenzene	0.561	0.561	0.2	0	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.