

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: Q1456 SAS No.: Q1456 SDG NO.: Q1456
 Lab File ID: VX045119.D Date Analyzed: 03/04/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	17184	4.89	86920	6.75	82480	10.05
UPPER LIMIT	34368	5.391	173840	7.251	164960	10.549
LOWER LIMIT	8592	4.391	43460	6.251	41240	9.549
EPA SAMPLE NO.						
EFF-WASTE WATER	15247	4.89	81841	6.76	77072	10.05
VX0304WBL01	14872	4.90	78579	6.76	69103	10.05
VX0304WBS01	16840	4.89	87021	6.76	81719	10.05
VX0304WBSD01	14793	4.89	81282	6.75	75372	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.