

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1462 SAS No.: Q1462 SDG No.: Q1462

Instrument ID: MSVOA_X Calibration Date/Time: 02/28/2025 10:32

Lab File ID: VX045077.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.770	0.741	0.1	-3.77	20
Vinyl Chloride	0.764	0.729		-4.58	20
Bromomethane	0.300	0.282		-6	20
Chloroethane	0.352	0.283		-19.6	20
Tert butyl alcohol	0.151	0.127		-15.89	20
1,1-Dichloroethene	0.614	0.584		-4.89	20
Acetone	0.369	0.312		-15.45	20
Carbon Disulfide	1.636	1.571		-3.97	20
Methyl tert-butyl Ether	2.061	1.935		-6.11	20
Methylene Chloride	0.731	0.667		-8.76	20
trans-1,2-Dichloroethene	0.607	0.586		-3.46	20
1,1-Dichloroethane	1.247	1.167	0.1	-6.41	20
2-Butanone	0.555	0.512		-7.75	20
Carbon Tetrachloride	0.465	0.459		-1.29	20
cis-1,2-Dichloroethene	0.749	0.712		-4.94	20
Chloroform	1.239	1.140		-7.99	20
1,1,1-Trichloroethane	1.000	0.948		-5.2	20
Benzene	1.448	1.433		-1.04	20
1,2-Dichloroethane	0.521	0.494		-5.18	20
Trichloroethene	0.340	0.328		-3.53	20
1,2-Dichloropropane	0.372	0.358		-3.76	20
Bromodichloromethane	0.516	0.508		-1.55	20
4-Methyl-2-Pentanone	0.587	0.565		-3.75	20
Toluene	0.849	0.870		2.47	20
t-1,3-Dichloropropene	0.431	0.483		12.06	20
cis-1,3-Dichloropropene	0.503	0.551		9.54	20
1,1,2-Trichloroethane	0.350	0.330		-5.71	20
2-Hexanone	0.425	0.411		-3.29	20
Dibromochloromethane	0.366	0.367		0.27	20
Tetrachloroethene	0.320	0.319		-0.31	20
Chlorobenzene	1.058	1.039	0.3	-1.8	20
Ethyl Benzene	1.843	1.879		1.95	20
m/p-Xylenes	0.675	0.694		2.82	20
o-Xylene	0.681	0.685		0.59	20
Styrene	1.101	1.143		3.82	20
Bromoform	0.266	0.268	0.1	0.75	20
1,1,2,2-Tetrachloroethane	1.432	1.334	0.3	-6.84	20
1,2-Dichloroethane-d4	0.795	0.733		-7.8	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1462 SAS No.: Q1462 SDG No.: Q1462

Instrument ID: MSVOA_X Calibration Date/Time: 02/28/2025 10:32

Lab File ID: VX045077.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dibromofluoromethane	0.334	0.331		-0.9	20
Toluene-d8	1.212	1.209		-0.25	20
4-Bromofluorobenzene	0.402	0.404		0.5	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.