

Surrogate Summary

SDG No.: **Q1462**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1462-01	MW2	1,2-Dichloroethane-d4	50	49.7	99	70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.8	114	70 (77)	130 (121)
Q1462-02	FB	1,2-Dichloroethane-d4	50	51.6	103	70 (74)	130 (125)
		Dibromofluoromethane	50	51.4	103	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.8	104	70 (77)	130 (121)
VX0228WBL01	VX0228WBL01	1,2-Dichloroethane-d4	50	50.0	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.0	102	70 (75)	130 (124)
		Toluene-d8	50	50.9	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.3	107	70 (77)	130 (121)
VX0228WBS01	VX0228WBS01	1,2-Dichloroethane-d4	50	46.9	94	70 (74)	130 (125)
		Dibromofluoromethane	50	48.2	96	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.0	102	70 (77)	130 (121)
VX0228WBSD0	VX0228WBSD01	1,2-Dichloroethane-d4	50	48.0	96	70 (74)	130 (125)
		Dibromofluoromethane	50	49.6	99	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.6	103	70 (77)	130 (121)

