

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1462

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX045081.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0228WBSD01	Chloromethane	20	18.6	ug/L	93	0		40 (65)	160 (116)	20 (20)
	Vinyl chloride	20	18.1	ug/L	91	1		70 (65)	130 (117)	20 (20)
	Bromomethane	20	18.7	ug/L	94	5		40 (58)	160 (125)	20 (20)
	Chloroethane	20	16.0	ug/L	80	4		40 (56)	160 (128)	20 (20)
	Tert butyl alcohol	100	84.8	ug/L	85	6		70 (73)	130 (124)	20 (20)
	1,1-Dichloroethene	20	18.2	ug/L	91	0		70 (74)	130 (110)	20 (20)
	Acetone	100	88.6	ug/L	89	7		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	17.9	ug/L	90	1		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	18.6	ug/L	93	3		70 (78)	130 (114)	20 (20)
	Methylene Chloride	20	17.9	ug/L	90	0		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	18.8	ug/L	94	1		70 (75)	130 (108)	20 (20)
	1,1-Dichloroethane	20	18.0	ug/L	90	0		70 (78)	130 (112)	20 (20)
	2-Butanone	100	95.9	ug/L	96	4		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	18.8	ug/L	94	1		70 (77)	130 (113)	20 (20)
	cis-1,2-Dichloroethene	20	18.9	ug/L	95	4		70 (77)	130 (110)	20 (20)
	Chloroform	20	18.5	ug/L	93	2		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	18.8	ug/L	94	3		70 (80)	130 (108)	20 (20)
	Benzene	20	19.3	ug/L	97	2		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	19.1	ug/L	96	3		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.0	ug/L	95	2		70 (77)	130 (113)	20 (20)
	1,2-Dichloropropane	20	19.0	ug/L	95	2		70 (83)	130 (111)	20 (20)
	Bromodichloromethane	20	19.0	ug/L	95	2		70 (83)	130 (110)	20 (20)
	4-Methyl-2-Pentanone	100	100	ug/L	100	4		40 (74)	160 (118)	20 (20)
	Toluene	20	19.9	ug/L	100	2		70 (82)	130 (110)	20 (20)
	t-1,3-Dichloropropene	20	19.9	ug/L	100	2		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	20.2	ug/L	101	1		70 (82)	130 (110)	20 (20)
	1,1,2-Trichloroethane	20	19.8	ug/L	99	3		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	100	ug/L	100	3		40 (73)	160 (117)	20 (20)
	Dibromochloromethane	20	19.2	ug/L	96	3		70 (82)	130 (110)	20 (20)
	Tetrachloroethene	20	19.7	ug/L	99	4		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.7	ug/L	99	1		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	19.8	ug/L	99	2		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	41.0	ug/L	103	3		70 (82)	130 (110)	20 (20)
	o-Xylene	20	19.8	ug/L	99	0		70 (83)	130 (109)	20 (20)
	Styrene	20	20.3	ug/L	102	1		70 (80)	130 (111)	20 (20)
	Bromoform	20	18.6	ug/L	93	2		70 (79)	130 (109)	20 (20)
	1,1,2,2-Tetrachloroethane	20	18.8	ug/L	94	2		70 (76)	130 (118)	20 (20)