

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1462 SAS No.: Q1462 SDG NO.: Q1462
 Lab File ID: VX045077.D Date Analyzed: 02/28/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:32
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	76314	12.018				
UPPER LIMIT	152628	12.518				
LOWER LIMIT	38157	11.518				
EPA SAMPLE NO.						
MW2	53843	12.02				
FB	54108	12.02				
VX0228WBL01	59690	12.02				
VX0228WBS01	76472	12.02				
VX0228WBSD01	73363	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.