

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1463 SAS No.: Q1463 SDG No.: Q1463

Instrument ID: MSVOA_Y Calibration Date/Time: 03/06/2025 10:10

Lab File ID: VY021426.D Init. Calib. Date(s): 03/04/2025 03/04/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 12:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.647	0.644	0.1	-0.46	20
Vinyl Chloride	0.707	0.731		3.39	20
Bromomethane	0.511	0.498		-2.54	20
Chloroethane	0.467	0.503		7.71	20
Tert butyl alcohol	0.041	0.037		-9.76	20
1,1-Dichloroethene	0.514	0.484		-5.84	20
Acetone	0.115	0.095		-17.39	20
Carbon Disulfide	1.610	1.552		-3.6	20
Methyl tert-butyl Ether	1.302	1.385		6.38	20
Methylene Chloride	0.605	0.568		-6.12	20
trans-1,2-Dichloroethene	0.575	0.563		-2.09	20
1,1-Dichloroethane	1.056	1.061	0.1	0.47	20
2-Butanone	0.159	0.160		0.63	20
Carbon Tetrachloride	0.592	0.577		-2.53	20
cis-1,2-Dichloroethene	0.653	0.658		0.77	20
Chloroform	1.096	1.107		1	20
1,1,1-Trichloroethane	1.004	0.973		-3.09	20
Benzene	1.515	1.544		1.91	20
1,2-Dichloroethane	0.424	0.447		5.43	20
Trichloroethene	0.383	0.378		-1.3	20
1,2-Dichloropropane	0.361	0.373		3.32	20
Bromodichloromethane	0.533	0.559		4.88	20
4-Methyl-2-Pentanone	0.228	0.266		16.67	20
Toluene	0.955	0.975		2.09	20
t-1,3-Dichloropropene	0.470	0.512		8.94	20
cis-1,3-Dichloropropene	0.557	0.587		5.39	20
1,1,2-Trichloroethane	0.263	0.285		8.36	20
2-Hexanone	0.151	0.180		19.2	20
Dibromochloromethane	0.360	0.391		8.61	20
Tetrachloroethene	0.407	0.384		-5.65	20
Chlorobenzene	1.181	1.160	0.3	-1.78	20
Ethyl Benzene	2.072	2.064		-0.39	20
m/p-Xylenes	0.787	0.784		-0.38	20
o-Xylene	0.725	0.733		1.1	20
Styrene	1.203	1.259		4.66	20
Bromoform	0.233	0.247	0.1	6.01	20
1,1,2,2-Tetrachloroethane	0.695	0.694	0.3	-0.14	20
1,2-Dichloroethane-d4	0.528	0.564		6.82	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1463 SAS No.: Q1463 SDG No.: Q1463

Instrument ID: MSVOA_Y Calibration Date/Time: 03/06/2025 10:10

Lab File ID: VY021426.D Init. Calib. Date(s): 03/04/2025 03/04/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 12:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dibromofluoromethane	0.329	0.349		6.08	20
Toluene-d8	1.244	1.310		5.3	20
4-Bromofluorobenzene	0.423	0.469		10.88	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.