

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1463 SAS No.: Q1463 SDG No.: Q1463

Instrument ID: MSVOA_Y Calibration Date/Time: 03/07/2025 09:59

Lab File ID: VY021451.D Init. Calib. Date(s): 03/04/2025 03/04/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 12:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.647	0.662	0.1	2.32	20
Vinyl Chloride	0.707	0.776		9.76	20
Bromomethane	0.511	0.511		0	20
Chloroethane	0.467	0.546		16.92	20
Tert butyl alcohol	0.041	0.030		-26.83	20
1,1-Dichloroethene	0.514	0.513		-0.19	20
Acetone	0.115	0.121		5.22	20
Carbon Disulfide	1.610	1.632		1.37	20
Methyl tert-butyl Ether	1.302	1.304		0.15	20
Methylene Chloride	0.605	0.575		-4.96	20
trans-1,2-Dichloroethene	0.575	0.587		2.09	20
1,1-Dichloroethane	1.056	1.095	0.1	3.69	20
2-Butanone	0.159	0.163		2.52	20
Carbon Tetrachloride	0.592	0.617		4.22	20
cis-1,2-Dichloroethene	0.653	0.675		3.37	20
Chloroform	1.096	1.152		5.11	20
1,1,1-Trichloroethane	1.004	1.033		2.89	20
Benzene	1.515	1.595		5.28	20
1,2-Dichloroethane	0.424	0.448		5.66	20
Trichloroethene	0.383	0.395		3.13	20
1,2-Dichloropropane	0.361	0.384		6.37	20
Bromodichloromethane	0.533	0.567		6.38	20
4-Methyl-2-Pentanone	0.228	0.236		3.51	20
Toluene	0.955	1.025		7.33	20
t-1,3-Dichloropropene	0.470	0.503		7.02	20
cis-1,3-Dichloropropene	0.557	0.592		6.28	20
1,1,2-Trichloroethane	0.263	0.278		5.7	20
2-Hexanone	0.151	0.163		7.95	20
Dibromochloromethane	0.360	0.381		5.83	20
Tetrachloroethene	0.407	0.413		1.47	20
Chlorobenzene	1.181	1.219	0.3	3.22	20
Ethyl Benzene	2.072	2.220		7.14	20
m/p-Xylenes	0.787	0.846		7.5	20
o-Xylene	0.725	0.771		6.34	20
Styrene	1.203	1.319		9.64	20
Bromoform	0.233	0.240	0.1	3	20
1,1,2,2-Tetrachloroethane	0.695	0.681	0.3	-2.01	20
1,2-Dichloroethane-d4	0.528	0.557		5.49	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1463 SAS No.: Q1463 SDG No.: Q1463

Instrument ID: MSVOA_Y Calibration Date/Time: 03/07/2025 09:59

Lab File ID: VY021451.D Init. Calib. Date(s): 03/04/2025 03/04/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 12:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dibromofluoromethane	0.329	0.352		6.99	20
Toluene-d8	1.244	1.351		8.6	20
4-Bromofluorobenzene	0.423	0.461		8.98	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.