

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1463 SAS No.: Q1463 SDG No.: Q1463

Instrument ID: MSVOA_Y Calibration Date/Time: 03/11/2025 09:42

Lab File ID: VY021474.D Init. Calib. Date(s): 03/04/2025 03/04/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 12:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
tert butyl alcohol	250.000	240.81	3.675201	20