

Report of Analysis

Client:	Holland Manufacturing Co.	Date Collected:	02/27/25 09:00
Project:	Pre Treatment Plant 2025	Date Received:	02/27/25
Client Sample ID:	CITY WATERDL	SDG No.:	Q1466
Lab Sample ID:	Q1466-05DL	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Chloride	19.6	D	5	0.055	3.00	mg/L		02/28/25 12:39	9056A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits