

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Holland Manufacturing Co. **SDG No.:** Q1466
Contract: HOLL01 **Lab Code:** CHEM **Case No.:** Q1466 **SAS No.:** Q1466
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Calcium	10000	10000	100	95 - 105	P	03/12/2025	10:52	LB135011
	Magnesium	5900	6000	98	95 - 105	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Holland Manufacturing Co. **SDG No.:** Q1466
Contract: HOLL01 **Lab Code:** CHEM **Case No.:** Q1466 **SAS No.:** Q1466
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Calcium	1960	2000	98	80 - 120	P	03/12/2025	10:56	LB135011
	Magnesium	2060	2000	103	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Holland Manufacturing Co. **SDG No.:** Q1466
Contract: HOLL01 **Lab Code:** CHEM **Case No.:** Q1466 **SAS No.:** Q1466
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Calcium	23800	25000	95	90 - 110	P	03/12/2025	11:29	LB135011
	Magnesium	23800	25000	95	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Calcium	23800	25000	95	90 - 110	P	03/12/2025	12:20	LB135011
	Magnesium	23900	25000	96	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Calcium	23800	25000	95	90 - 110	P	03/12/2025	13:10	LB135011
	Magnesium	24000	25000	96	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Calcium	23800	25000	95	90 - 110	P	03/12/2025	14:24	LB135011
	Magnesium	24000	25000	96	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Calcium	23500	25000	94	90 - 110	P	03/12/2025	15:25	LB135011
	Magnesium	23900	25000	96	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Calcium	23200	25000	93	90 - 110	P	03/12/2025	16:25	LB135011
	Magnesium	23700	25000	95	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Calcium	23100	25000	92	90 - 110	P	03/12/2025	17:25	LB135011
	Magnesium	23400	25000	93	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Calcium	23600	25000	94	90 - 110	P	03/12/2025	18:23	LB135011
	Magnesium	24300	25000	97	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Calcium	23600	25000	95	90 - 110	P	03/12/2025	18:49	LB135011
	Magnesium	24100	25000	96	90 - 110	P	03/12/2025	18:49	LB135011