

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Holland Manufacturing Co. **SDG No.:** Q1466
Contract: HOLL01 **Lab Code:** CHEM **Case No.:** Q1466 **SAS No.:** Q1466
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Calcium	232000	245000	95	208000	282000	03/12/2025	11:09	LB135011
	Magnesium	249000	255000	98	216000	294000	03/12/2025	11:09	LB135011
ICSAB01	Calcium	220000	235000	94	199000	271000	03/12/2025	11:14	LB135011
	Magnesium	237000	248000	96	210000	286000	03/12/2025	11:14	LB135011
ICSA	Calcium	238000	245000	97	208000	282000	03/12/2025	11:21	LB135011
	Magnesium	245000	255000	96	216000	294000	03/12/2025	11:21	LB135011
ICSAB	Calcium	240000	235000	102	199000	271000	03/12/2025	11:25	LB135011
	Magnesium	247000	248000	100	210000	286000	03/12/2025	11:25	LB135011