

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Holland Manufacturing Co. **level:** low **sdg no.:** Q1466
contract: HOLL01 **lab code:** CHEM **case no.:** Q1466 **sas no.:** Q1466
matrix: Water **sample id:** Q1466-05 **client id:** CITY WATERMS
Percent Solids for Sample: NA **Spiked ID:** Q1466-05MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Calcium	ug/L	75 - 125	22500		22000		500	101		P
Magnesium	ug/L	75 - 125	12100		11300		1000	86		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Holland Manufacturing Co. **level:** low **sdg no.:** Q1466
contract: HOLL01 **lab code:** CHEM **case no.:** Q1466 **sas no.:** Q1466
matrix: Water **sample id:** Q1466-05 **client id:** CITY WATERMSD
Percent Solids for Sample: NA **Spiked ID:** Q1466-05MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Calcium	ug/L	75 - 125	22400		22000		500	73		P
Magnesium	ug/L	75 - 125	12100		11300		1000	79		P