

Initial and Continuing Calibration Verification

Client: American Wax

SDG No.: Q1468

Project: Semi Annual 2025

RunNo.: LB134926

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: COD	CCV1	mg/L	50.319	50	101	95-105	03/10/2025
Sample ID: COD	CCV2	mg/L	48.339	50	97	95-105	03/10/2025