

Laboratory Control Sample Summary

Client: American Wax

SDG No.: Q1468

Project: Semi Annual 2025

Run No.: LB134868

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134868BS							
BOD_7	mg/L	198	186		94	1	84.6-115.4	02/28/2025

Laboratory Control Sample Summary

Client: American Wax

SDG No.: Q1468

Project: Semi Annual 2025

Run No.: LB134882

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134882BS							
Non-Polar Material	mg/L	20.0	16.7		84	1	78-114	03/03/2025

Laboratory Control Sample Summary

Client: American Wax

SDG No.: Q1468

Project: Semi Annual 2025

Run No.: LB134882

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134882BSD							
Non-Polar Material	mg/L	20.0	17.2		86	1	78-114	03/03/2025

Laboratory Control Sample Summary

Client: American Wax

SDG No.: Q1468

Project: Semi Annual 2025

Run No.: LB134926

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134926BS							
COD	mg/L	50	49.3		99	1	90-110	03/10/2025