

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1469 SAS No.: Q1469 SDG NO.: Q1469

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/04/2025

Lab Sample No.: INDA319 Data File : PD087947.D Time Analyzed: 13:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.639	6.779	2568240.000	2400510.000	-6.5
4,4'-DDT	7.025	6.955	7.095	2665660.000	2358420.000	-11.5
alpha-BHC	4.003	3.953	4.053	4131350.000	3608980.000	-12.6
DCB	9.080	8.980	9.180	3294960.000	2661280.000	-19.2
Dieldrin	6.352	6.282	6.422	3644250.000	3331160.000	-8.6
Endosulfan I	6.079	6.009	6.149	3360700.000	3011420.000	-10.4
Endrin	6.579	6.509	6.649	3028440.000	2835130.000	-6.4
gamma-BHC (Lindane)	4.334	4.284	4.384	4101440.000	3562010.000	-13.2
Heptachlor	4.934	4.884	4.984	4025040.000	3596370.000	-10.7
Methoxychlor	7.497	7.427	7.567	1464130.000	1320020.000	-9.8
TCX	3.553	3.503	3.603	2086660.000	1767870.000	-15.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1469 SAS No.: Q1469 SDG NO.: Q1469

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/04/2025

Lab Sample No.: INDA319 Data File : PD087947.D Time Analyzed: 13:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.936	5.866	6.006	14429100.000	14716500.000	2.0
4,4'-DDT	6.190	6.120	6.260	14643200.000	15335100.000	4.7
alpha-BHC	3.397	3.347	3.447	21057100.000	21796300.000	3.5
DCB	8.080	7.980	8.180	15809900.000	15793400.000	-0.1
Dieldrin	5.519	5.449	5.589	17907700.000	18437300.000	3.0
Endosulfan I	5.253	5.183	5.323	16918600.000	16460600.000	-2.7
Endrin	5.795	5.725	5.865	15984600.000	16335100.000	2.2
gamma-BHC (Lindane)	3.734	3.684	3.784	19255900.000	20084900.000	4.3
Heptachlor	4.088	4.038	4.138	19870200.000	20453800.000	2.9
Methoxychlor	6.761	6.691	6.831	6618460.000	6561740.000	-0.9
TCX	2.884	2.834	2.934	13450900.000	13817100.000	2.7