

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1469 SAS No.: Q1469 SDG NO.: Q1469

EPA Sample No.: SSTD020625 Date Analyzed: 03/04/2025

Lab File ID: BP024146.D Time Analyzed: 10:59

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	330869	7.758	1369680	10.53	835568	14.39
UPPER LIMIT	661738	8.258	2739360	11.034	1671140	14.892
LOWER LIMIT	165435	7.258	684840	10.034	417784	13.892
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	231480	7.75	918724	10.53	586861	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.