

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1469

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VX045081.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0228WBSD01	Dichlorodifluoromethane	20	18.9	ug/L	95	1		69	116	20
	Chloromethane	20	18.6	ug/L	93	0		65	116	20
	Vinyl chloride	20	18.1	ug/L	91	1		65	117	20
	Ethyl Acetate	20	19.0	ug/L	95	7		75	115	20
	Bromomethane	20	18.7	ug/L	94	5		58	125	20
	Chloroethane	20	16.0	ug/L	80	4		56	128	20
	Trichlorofluoromethane	20	19.1	ug/L	96	3		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	20.1	ug/L	101	1		80	112	20
	Tert butyl alcohol	100	84.8	ug/L	85	6		73	124	20
	1,1-Dichloroethene	20	18.2	ug/L	91	0		74	110	20
	Acrolein	100	120	ug/L	120	9		51	143	20
	Acrylonitrile	100	93.9	ug/L	94	5		73	113	20
	Acetone	100	88.6	ug/L	89	7		60	125	20
	Carbon disulfide	20	17.9	ug/L	90	1		64	112	20
	Methyl tert-butyl Ether	20	18.6	ug/L	93	3		78	114	20
	Methyl Acetate	20	18.5	ug/L	93	6		67	125	20
	Methylene Chloride	20	17.9	ug/L	90	0		72	114	20
	trans-1,2-Dichloroethene	20	18.8	ug/L	94	1		75	108	20
	Vinyl Acetate	100	98.3	ug/L	98	4		76	115	20
	1,1-Dichloroethane	20	18.0	ug/L	90	0		78	112	20
	Cyclohexane	20	19.6	ug/L	98	1		75	110	20
	2-Butanone	100	95.9	ug/L	96	4		65	122	20
	Carbon Tetrachloride	20	18.8	ug/L	94	1		77	113	20
	2,2-Dichloropropane	20	26.8	ug/L	134	2	*	75	116	20
	cis-1,2-Dichloroethene	20	18.9	ug/L	95	4		77	110	20
	Bromochloromethane	20	18.1	ug/L	91	4		70	124	20
	Chloroform	20	18.5	ug/L	93	2		79	113	20
	1,1,1-Trichloroethane	20	18.8	ug/L	94	3		80	108	20
	Methylcyclohexane	20	21.5	ug/L	108	4		72	115	20
	1,1-Dichloropropene	20	18.5	ug/L	93	1		79	110	20
	Benzene	20	19.3	ug/L	97	2		82	109	20
	1,2-Dichloroethane	20	19.1	ug/L	96	3		80	115	20
	Trichloroethene	20	19.0	ug/L	95	2		77	113	20
	1,2-Dichloropropane	20	19.0	ug/L	95	2		83	111	20
	Dibromomethane	20	19.4	ug/L	97	2		82	110	20
	Bromodichloromethane	20	19.0	ug/L	95	2		83	110	20
	4-Methyl-2-Pentanone	100	100	ug/L	100	4		74	118	20
	Toluene	20	19.9	ug/L	100	2		82	110	20
	t-1,3-Dichloropropene	20	19.9	ug/L	100	2		79	110	20
	cis-1,3-Dichloropropene	20	20.2	ug/L	101	1		82	110	20
	1,1,2-Trichloroethane	20	19.8	ug/L	99	3		83	112	20
	1,3-Dichloropropane	20	19.6	ug/L	98	3		83	111	20
	2-Chloroethyl vinyl ether	100	100	ug/L	100	0		75	124	20
	2-Hexanone	100	100	ug/L	100	3		73	117	20
	Dibromochloromethane	20	19.2	ug/L	96	3		82	110	20
	1,2-Dibromoethane	20	19.8	ug/L	99	2		81	110	20
	Tetrachloroethene	20	19.7	ug/L	99	4		67	123	20
	Chlorobenzene	20	19.7	ug/L	99	1		82	109	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1469

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Datafile : VX045081.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0228WBSD01	1,1,1,2-Tetrachloroethane	20	18.8	ug/L	94	2		84	111	20
	Hexachloroethane	20	18.2	ug/L	91	2		80	113	20
	Ethyl Benzene	20	19.8	ug/L	99	2		83	109	20
	m/p-Xylenes	40	41.0	ug/L	103	3		82	110	20
	o-Xylene	20	19.8	ug/L	99	0		83	109	20
	Styrene	20	20.3	ug/L	102	1		80	111	20
	Bromoform	20	18.6	ug/L	93	2		79	109	20
	Isopropylbenzene	20	19.7	ug/L	99	2		83	112	20
	1,1,2,2-Tetrachloroethane	20	18.8	ug/L	94	2		76	118	20
	1,2,3-Trichloropropane	20	18.7	ug/L	94	5		75	112	20
	Bromobenzene	20	19.2	ug/L	96	0		77	116	20
	N-propylbenzene	20	19.9	ug/L	100	0		83	112	20
	2-Chlorotoluene	20	19.1	ug/L	96	1		83	110	20
	1,3,5-Trimethylbenzene	20	20.1	ug/L	101	0		85	112	20
	4-Chlorotoluene	20	19.9	ug/L	100	3		82	110	20
	tert-Butylbenzene	20	19.3	ug/L	97	0		83	112	20
	1,2,4-Trimethylbenzene	20	19.9	ug/L	100	2		85	111	20
	Sec-butylbenzene	20	20.3	ug/L	102	2		81	114	20
	p-Isopropyltoluene	20	20.3	ug/L	102	0		78	116	20
	1,3-Dichlorobenzene	20	19.5	ug/L	98	0		82	108	20
	1,4-Dichlorobenzene	20	19.5	ug/L	98	1		82	107	20
	n-Butylbenzene	20	20.4	ug/L	102	0		75	115	20
	1,2-Dichlorobenzene	20	19.8	ug/L	99	2		82	109	20
	1,2-Dibromo-3-Chloropropane	20	18.5	ug/L	93	4		68	112	20
	1,2,4-Trichlorobenzene	20	20.7	ug/L	104	5		75	113	20
	Hexachlorobutadiene	20	20.4	ug/L	102	0		81	121	20
	Naphthalene	20	20.1	ug/L	101	5		78	119	20
	1,2,3-Trichlorobenzene	20	20.4	ug/L	102	2		76	114	20