

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02  
 Lab Code: CHEM Case No.: Q1469 SAS No.: Q1469 SDG NO.: Q1469  
 Lab File ID: VX045077.D Date Analyzed: 02/28/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 10:32  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                                 | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|---------------------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD                     | 111989        | 5.54  | 196414        | 6.75  | 172369        | 10.05  |
| UPPER LIMIT                     | 223978        | 6.037 | 392828        | 7.251 | 344738        | 10.549 |
| LOWER LIMIT                     | 55994.5       | 5.037 | 98207         | 6.251 | 86184.5       | 9.549  |
| EPA SAMPLE NO.                  |               |       |               |       |               |        |
| Storage-Blank-SOIL-REF-6        | 71717         | 5.55  | 141247        | 6.76  | 130502        | 10.06  |
| Storage-Blank-WATER-REF-5       | 67134         | 5.54  | 133806        | 6.76  | 124730        | 10.05  |
| Storage-Blank-WATER-REF-4       | 78776         | 5.54  | 154922        | 6.76  | 143838        | 10.05  |
| Storage-Blank-SAMPLE-MANAGEMENT | 69889         | 5.54  | 139482        | 6.76  | 127529        | 10.05  |
| VX0228WBL01                     | 76430         | 5.54  | 149381        | 6.76  | 137180        | 10.05  |
| VX0228WBS01                     | 108329        | 5.54  | 192724        | 6.76  | 169474        | 10.05  |
| VX0228WBSD01                    | 102676        | 5.54  | 183348        | 6.76  | 162382        | 10.05  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.