

LAB CHRONICLE

OrderID:	Q1477	OrderDate:	3/3/2025 10:23:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	CTO WE13
Contact:	Ernie Wu	Location:	H41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1477-01	RW7-SP100-20250228	Water			02/28/25			03/03/25
			Metals Group4	6010D		03/04/25	03/12/25	
Q1477-04	RW7-SP303-20250228	Water			02/28/25			03/03/25
			Metals Group4	6010D		03/04/25	03/12/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1477

Order ID: Q1477

Client: Tetra Tech NUS, Inc.

Project ID: CTO WE13

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : RW7-SP100-20250228									
Q1477-01	RW7-SP100-20250228	Water	Iron	248		18.5	40.0	50.0	ug/L
Client ID : RW7-SP303-20250228									
Q1477-04	RW7-SP303-20250228	Water	Iron	32.2	J	18.5	40.0	50.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/28/25
Project:	CTO WE13	Date Received:	03/03/25
Client Sample ID:	RW7-SP100-20250228	SDG No.:	Q1477
Lab Sample ID:	Q1477-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	248		1	18.5	40.0	50.0	ug/L	03/04/25 09:10	03/12/25 18:40	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/28/25
Project:	CTO WE13	Date Received:	03/03/25
Client Sample ID:	RW7-SP303-20250228	SDG No.:	Q1477
Lab Sample ID:	Q1477-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	32.2	J	1	18.5	40.0	50.0	ug/L	03/04/25 09:10	03/12/25 18:44	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	10400	10000	104	90 - 110	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	114	100	114	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4790	5000	96	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Iron	4890	5000	98	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Iron	4740	5000	95	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Iron	4970	5000	100	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Iron	4840	5000	97	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Iron	4770	5000	96	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Iron	4870	5000	97	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Iron	4810	5000	96	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Iron	4960	5000	99	90 - 110	P	03/12/2025	18:49	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	9860	10000	99	90 - 110	P	03/13/2025	12:07	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	101	100	101	80 - 120	P	03/13/2025	12:13	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4890	5000	98	90 - 110	P	03/13/2025	12:53	LB135035
CCV02	Iron	5130	5000	103	90 - 110	P	03/13/2025	13:49	LB135035
CCV03	Iron	4990	5000	100	90 - 110	P	03/13/2025	14:40	LB135035
CCV04	Iron	4780	5000	96	90 - 110	P	03/13/2025	15:38	LB135035
CCV05	Iron	5120	5000	102	90 - 110	P	03/13/2025	16:28	LB135035
CCV06	Iron	5040	5000	101	90 - 110	P	03/13/2025	17:21	LB135035
CCV07	Iron	5150	5000	103	90 - 110	P	03/13/2025	18:12	LB135035
CCV08	Iron	4880	5000	98	90 - 110	P	03/13/2025	18:53	LB135035
CCV09	Iron	5230	5000	105	90 - 110	P	03/13/2025	19:12	LB135035



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	98.7	100	99	40 - 160	P	03/12/2025	11:05	LB135011
CRI01	Iron	110	100	110	40 - 160	P	03/13/2025	12:21	LB135035



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tetra Tech NUS, Inc.				SDG No.:	Q1477				
Contract:	TETR06	Lab Code:	CHEM	Case No.:	Q1477	SAS No.:	Q1477			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	100	+/-100	U	80.0	100	P	03/12/2025	11:01	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	100	+/-100	U	80.0	100	P	03/12/2025	11:34	LB135011
CCB02	Iron	100	+/-100	U	80.0	100	P	03/12/2025	12:24	LB135011
CCB03	Iron	100	+/-100	U	80.0	100	P	03/12/2025	13:14	LB135011
CCB04	Iron	100	+/-100	U	80.0	100	P	03/12/2025	14:28	LB135011
CCB05	Iron	100	+/-100	U	80.0	100	P	03/12/2025	15:29	LB135011
CCB06	Iron	100	+/-100	U	80.0	100	P	03/12/2025	16:30	LB135011
CCB07	Iron	100	+/-100	U	80.0	100	P	03/12/2025	17:30	LB135011
CCB08	Iron	100	+/-100	U	80.0	100	P	03/12/2025	18:27	LB135011
CCB09	Iron	100	+/-100	U	80.0	100	P	03/12/2025	18:53	LB135011

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	100	+/-100	U	80.0	100	P	03/13/2025	12:17	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	100	+/-100	U	80.0	100	P	03/13/2025	12:58	LB135035
CCB02	Iron	100	+/-100	U	80.0	100	P	03/13/2025	13:54	LB135035
CCB03	Iron	100	+/-100	U	80.0	100	P	03/13/2025	14:48	LB135035
CCB04	Iron	100	+/-100	U	80.0	100	P	03/13/2025	15:42	LB135035
CCB05	Iron	100	+/-100	U	80.0	100	P	03/13/2025	16:32	LB135035
CCB06	Iron	100	+/-100	U	80.0	100	P	03/13/2025	17:25	LB135035
CCB07	Iron	45.6	+/-100	J	80.0	100	P	03/13/2025	18:35	LB135035
CCB08	Iron	100	+/-100	U	80.0	100	P	03/13/2025	19:03	LB135035
CCB09	Iron	100	+/-100	U	80.0	100	P	03/13/2025	19:21	LB135035

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166963BL		WATER		Batch Number:		PB166963		Prep Date:	03/04/2025	
	Iron	50.0	<50.0	U	40.0	50.0	P	03/13/2025	13:02	LB135035

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Tetra Tech NUS, Inc.</u>	SDG No.: <u>Q1477</u>
Contract: <u>TETR06</u> Lab Code: <u>CHEM</u>	Case No.: <u>Q1477</u> SAS No.: <u>Q1477</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	104000	101000	103	85600	116500	03/12/2025	11:09	LB135011
ICSAB01	Iron	102000	99300	103	84400	114500	03/12/2025	11:14	LB135011
ICSA01	Iron	103000	101000	102	85600	116500	03/13/2025	12:25	LB135035
ICSAB01	Iron	100000	99300	101	84400	114500	03/13/2025	12:30	LB135035



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tetra Tech NUS, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q1477</u>
contract: <u>TETR06</u>	lab code: <u>CHEM</u>	case no.: <u>Q1477</u> sas no.: <u>Q1477</u>
matrix: <u>Water</u>	sample id: <u>Q1478-07</u>	client id: <u>IDW-AQ-IW-03-COMP-022825MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1478-07MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	5430		3820		1500	108		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Tetra Tech NUS, Inc. level: low sdg no.: Q1477
contract: TETR06 lab code: CHEM case no.: Q1477 sas no.: Q1477
matrix: Water sample id: Q1478-07 client id: IDW-AQ-IW-03-COMP-022825MSD
Percent Solids for Sample: NA Spiked ID: Q1478-07MSD Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	5120		3820		1500	87		P

Metals
- 5b -

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Contract: TETR06

Lab Code: CHEM

Case No.: Q1477

SAS No.: Q1477

Matrix:

Level: LOW

Client ID:

Sample ID:

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1477</u>
Contract:	<u>TETR06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1477</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1478-07</u>	Client ID:	<u>IDW-AQ-IW-03-COMP-022825DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1478-07DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	3820		3940		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1477</u>
Contract:	<u>TETR06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1477</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1478-07MS</u>	Client ID:	<u>IDW-AQ-IW-03-COMP-022825MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1478-07MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	5430		5120		6		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166963BS Iron	ug/L	1500	1420		95	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

IDW-AQ-IW-03-COMP-022825L

Lab Name: Chemtech Consulting Group

Contract: TETR06

Lab Code: CHEM

Lb No.: lb135011

Lab Sample ID : Q1478-07L

SDG No.: Q1477

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Iron	3820	3800	0		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Contract: TETR06

Lab Code: CHEM

Case No.: Q1477

SAS No.: Q1477

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Contract: TETR06

Lab Code: CHEM

Case No.: Q1477

SAS No.: Q1477

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Contract: TETR06

Lab Code: CHEM

Case No.: Q1477

SAS No.: Q1477

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Contract: TETR06

Lab Code: CHEM

Case No.: Q1477

SAS No.: Q1477

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q1477

Contract: TETR06

Lab Code: CHEM

Case No.: Q1477

SAS No.: Q1477

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL
PREPARATION &
ANALYICAL
SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	SDG No.:	<u>Q1477</u>
Contract:	<u>TETR06</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>Q1477</u>
		SAS No.:	<u>Q1477</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB166963							
PB166963BL	PB166963BL	MB	WATER	03/04/2025	50.0	25.0	
PB166963BS	PB166963BS	LCS	WATER	03/04/2025	50.0	25.0	
Q1477-01	RW7-SP100-20250228	SAM	WATER	03/04/2025	50.0	25.0	
Q1477-04	RW7-SP303-20250228	SAM	WATER	03/04/2025	50.0	25.0	
Q1478-07DUP	IDW-AQ-IW-03-COMP-022825DUP	DUP	WATER	03/04/2025	50.0	25.0	
Q1478-07MS	IDW-AQ-IW-03-COMP-022825MS	MS	WATER	03/04/2025	50.0	25.0	
Q1478-07MSD	IDW-AQ-IW-03-COMP-022825MSD	MSD	WATER	03/04/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Tetra Tech NUS, Inc. **Contract:** TETR06

Lab code: CHEM **Case no.:** Q1477 **Sas no.:** Q1477 **Sdg no.:** Q1477

Instrument id number: **Method:** **Run number:** LB135011

Start date: 03/12/2025 **End date:** 03/12/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1027	Fe
S1	S1	1	1031	Fe
S2	S2	1	1035	Fe
S3	S3	1	1039	Fe
S4	S4	1	1044	Fe
S5	S5	1	1048	Fe
ICV01	ICV01	1	1052	Fe
LLICV01	LLICV01	1	1056	Fe
ICB01	ICB01	1	1101	Fe
CRI01	CRI01	1	1105	Fe
ICSA01	ICSA01	1	1109	Fe
ICSAB01	ICSAB01	1	1114	Fe
CCV01	CCV01	1	1129	Fe
CCB01	CCB01	1	1134	Fe
CCV02	CCV02	1	1220	Fe
CCB02	CCB02	1	1224	Fe
CCV03	CCV03	1	1310	Fe
CCB03	CCB03	1	1314	Fe
CCV04	CCV04	1	1424	Fe
CCB04	CCB04	1	1428	Fe
CCV05	CCV05	1	1525	Fe
CCB05	CCB05	1	1529	Fe
CCV06	CCV06	1	1625	Fe
CCB06	CCB06	1	1630	Fe
CCV07	CCV07	1	1725	Fe
CCB07	CCB07	1	1730	Fe
Q1478-07DUP	IDW-AQ-IW-03-COMP-022825	1	1752	Fe
Q1478-07L	IDW-AQ-IW-03-COMP-022825	5	1756	Fe
Q1478-07MS	IDW-AQ-IW-03-COMP-022825	1	1800	Fe
Q1478-07MSD	IDW-AQ-IW-03-COMP-022825	1	1805	Fe
CCV08	CCV08	1	1823	Fe
CCB08	CCB08	1	1827	Fe
Q1477-01	RW7-SP100-20250228	1	1840	Fe
Q1477-04	RW7-SP303-20250228	1	1844	Fe
CCV09	CCV09	1	1849	Fe
CCB09	CCB09	1	1853	Fe

metals
- 14 -
ANALYSIS RUN LOG

Client: Tetra Tech NUS, Inc. **Contract:** TETR06

Lab code: CHEM **Case no.:** Q1477 **Sas no.:** Q1477 **Sdg no.:** Q1477

Instrument id number: **Method:** **Run number:** LB135035

Start date: 03/13/2025 **End date:** 03/13/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1142	Fe
S1	S1	1	1146	Fe
S2	S2	1	1150	Fe
S3	S3	1	1154	Fe
S4	S4	1	1159	Fe
S5	S5	1	1203	Fe
ICV01	ICV01	1	1207	Fe
LLICV01	LLICV01	1	1213	Fe
ICB01	ICB01	1	1217	Fe
CRI01	CRI01	1	1221	Fe
ICSA01	ICSA01	1	1225	Fe
ICSAB01	ICSAB01	1	1230	Fe
CCV01	CCV01	1	1253	Fe
CCB01	CCB01	1	1258	Fe
PB166963BL	PB166963BL	1	1302	Fe
PB166963BS	PB166963BS	1	1306	Fe
CCV02	CCV02	1	1349	Fe
CCB02	CCB02	1	1354	Fe
CCV03	CCV03	1	1440	Fe
CCB03	CCB03	1	1448	Fe
CCV04	CCV04	1	1538	Fe
CCB04	CCB04	1	1542	Fe
CCV05	CCV05	1	1628	Fe
CCB05	CCB05	1	1632	Fe
CCV06	CCV06	1	1721	Fe
CCB06	CCB06	1	1725	Fe
CCV07	CCV07	1	1812	Fe
CCB07	CCB07	1	1835	Fe
CCV08	CCV08	1	1853	Fe
CCB08	CCB08	1	1903	Fe
CCV09	CCV09	1	1912	Fe
CCB09	CCB09	1	1921	Fe