

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	10400	10000	104	90 - 110	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	114	100	114	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4790	5000	96	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Iron	4890	5000	98	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Iron	4740	5000	95	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Iron	4970	5000	100	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Iron	4840	5000	97	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Iron	4770	5000	96	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Iron	4870	5000	97	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Iron	4810	5000	96	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Iron	4960	5000	99	90 - 110	P	03/12/2025	18:49	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	9860	10000	99	90 - 110	P	03/13/2025	12:07	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	101	100	101	80 - 120	P	03/13/2025	12:13	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4890	5000	98	90 - 110	P	03/13/2025	12:53	LB135035
CCV02	Iron	5130	5000	103	90 - 110	P	03/13/2025	13:49	LB135035
CCV03	Iron	4990	5000	100	90 - 110	P	03/13/2025	14:40	LB135035
CCV04	Iron	4780	5000	96	90 - 110	P	03/13/2025	15:38	LB135035
CCV05	Iron	5120	5000	102	90 - 110	P	03/13/2025	16:28	LB135035
CCV06	Iron	5040	5000	101	90 - 110	P	03/13/2025	17:21	LB135035
CCV07	Iron	5150	5000	103	90 - 110	P	03/13/2025	18:12	LB135035
CCV08	Iron	4880	5000	98	90 - 110	P	03/13/2025	18:53	LB135035
CCV09	Iron	5230	5000	105	90 - 110	P	03/13/2025	19:12	LB135035