



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	98.7	100	99	40 - 160	P	03/12/2025	11:05	LB135011
CRI01	Iron	110	100	110	40 - 160	P	03/13/2025	12:21	LB135035