



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Tetra Tech NUS, Inc. **SDG No.:** Q1477  
**Contract:** TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| ICB01     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 11:01            | LB135011      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Tetra Tech NUS, Inc. **SDG No.:** Q1477  
**Contract:** TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| CCB01     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 11:34            | LB135011      |
| CCB02     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 12:24            | LB135011      |
| CCB03     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 13:14            | LB135011      |
| CCB04     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 14:28            | LB135011      |
| CCB05     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 15:29            | LB135011      |
| CCB06     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 16:30            | LB135011      |
| CCB07     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 17:30            | LB135011      |
| CCB08     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 18:27            | LB135011      |
| CCB09     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/12/2025       | 18:53            | LB135011      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Tetra Tech NUS, Inc.

**SDG No.:** Q1477

**Contract:** TETR06

**Lab Code:** CHEM

**Case No.:** Q1477

**SAS No.:** Q1477

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| ICB01     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 12:17            | LB135035      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. SDG No.: Q1477  
Contract: TETR06 Lab Code: CHEM Case No.: Q1477 SAS No.: Q1477

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| CCB01     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 12:58            | LB135035      |
| CCB02     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 13:54            | LB135035      |
| CCB03     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 14:48            | LB135035      |
| CCB04     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 15:42            | LB135035      |
| CCB05     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 16:32            | LB135035      |
| CCB06     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 17:25            | LB135035      |
| CCB07     | Iron    | 45.6           | +/-100              | J            | 80.0 | 100  | P | 03/13/2025       | 18:35            | LB135035      |
| CCB08     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 19:03            | LB135035      |
| CCB09     | Iron    | 100            | +/-100              | U            | 80.0 | 100  | P | 03/13/2025       | 19:21            | LB135035      |