

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1477
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1477 **SAS No.:** Q1477
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	104000	101000	103	85600	116500	03/12/2025	11:09	LB135011
ICSAB01	Iron	102000	99300	103	84400	114500	03/12/2025	11:14	LB135011
ICSA01	Iron	103000	101000	102	85600	116500	03/13/2025	12:25	LB135035
ICSAB01	Iron	100000	99300	101	84400	114500	03/13/2025	12:30	LB135035