

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q1477
contract: TETR06 **lab code:** CHEM **case no.:** Q1477 **sas no.:** Q1477
matrix: Water **sample id:** Q1478-07 **client id:** IDW-AQ-IW-03-COMP-022825MS
Percent Solids for Sample: NA **Spiked ID:** Q1478-07MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	5430		3820		1500	108		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q1477
contract: TETR06 **lab code:** CHEM **case no.:** Q1477 **sas no.:** Q1477
matrix: Water **sample id:** Q1478-07 **client id:** IDW-AQ-IW-03-COMP-022825MSD
Percent Solids for Sample: NA **Spiked ID:** Q1478-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	5120		3820		1500	87		P