



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger STC PTC Site # D3868221**

**Project # N/A**

**Chemtech Project # Q1478**

**Test Name: Corrosivity,pH,Flash Point,Ignitability**

### **A. Number of Samples and Date of Receipt:**

4 Solid samples were received on 02/28/2025.

8 Water samples were received on 02/28/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Flash Point, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-RCRA, METALS RCRA, PCB, pH, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA, TCLP ZHE Extraction and VOC-TCLVOA-10. This data package contains results for Corrosivity,pH,Flash Point,Ignitability.

### **C. Analytical Techniques:**

The analysis of Flash Point was based on method 1010B, The analysis of Ignitability was based on method 1030, The analysis of pH was based on method 9040C and The analysis of Corrosivity was based on method 9045D.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for IDW-AQ-IW-01-COMP-022825 of pH, for IDW-AQ-IW-02-COMP-022825 of pH.for IDW-AQ-IW-03-COMP-022825 of pH.for IDW-AQ-MW-19B-COMP-022825 of pH.for IDW-SO-COMP-022825 of Corrosivity. As these samples are received out of hold.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_