



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger STC PTC Site # D3868221**

**Project # N/A**

**Chemtech Project # Q1478**

**Test Name: Metals ICP-RCRA,Mercury**

### **A. Number of Samples and Date of Receipt:**

4 Solid samples were received on 02/28/2025.

8 Water samples were received on 02/28/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Flash Point, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-RCRA, METALS RCRA, PCB, pH, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA, TCLP ZHE Extraction and VOC-TCLVOA-10. This data package contains results for Metals ICP-RCRA,Mercury.

### **C. Analytical Techniques:**

The analysis of Metals ICP-RCRA was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of Mercury was based on method 7470A.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (IDW-AQ-IW-03-COMP-022825MS) analysis met criteria for all samples except for Silver.

The Matrix Spike Duplicate (IDW-AQ-IW-03-COMP-022825MSD) analysis met criteria for all samples except for Barium, Silver.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.



Signature\_\_\_\_\_