



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1478

Project: Former Schlumberger STC PTC Site # D3868221

RunNo.: LB134873

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	03/03/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	03/03/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	03/03/2025
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	03/03/2025

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1478

Project: Former Schlumberger STC PTC Site # D3868221

RunNo.: LB134900

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	03/04/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	03/04/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	03/04/2025

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1478

Project: Former Schlumberger STC PTC Site # D3868221

RunNo.: LB134999

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.4	81	102	78-84	03/12/2025