

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134958

Review By	Mohan	Review On	3/11/2025 9:32:51 AM
Supervise By	jaswal	Supervise On	3/11/2025 9:38:08 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84793,MP84794,MP84795,MP84796,MP84797,MP84798		
ICV Standard	MP84799		
CCV Standard	MP84801		
ICSA Standard			
CRI Standard	MP84803		
LCS Standard			
Chk Standard	MP84800,MP84802,MP84804,MP84826		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/10/25 09:59		Mohan	OK
2	S0.2	S0.2	CAL2	03/10/25 10:01		Mohan	OK
3	S2.5	S2.5	CAL3	03/10/25 10:03		Mohan	OK
4	S5	S5	CAL4	03/10/25 10:06		Mohan	OK
5	S7.5	S7.5	CAL5	03/10/25 10:13		Mohan	OK
6	S10	S10	CAL6	03/10/25 10:19		Mohan	OK
7	ICV35	ICV35	ICV	03/10/25 10:22		Mohan	OK
8	ICB35	ICB35	ICB	03/10/25 10:26		Mohan	OK
9	CCV75	CCV75	CCV	03/10/25 10:34		Mohan	OK
10	CCB75	CCB75	CCB	03/10/25 10:39		Mohan	OK
11	CRA	CRA	CRDL	03/10/25 10:41		Mohan	OK
12	HighStd	HighStd	HIGH STD	03/10/25 10:43		Mohan	OK
13	ChkStd	ChkStd	SAM	03/10/25 10:46		Mohan	OK
14	PB167033BL	PB167033BL	MB	03/10/25 10:53		Mohan	OK
15	PB167033BS	PB167033BS	LCS	03/10/25 10:56		Mohan	OK
16	Q1488-02	ENV-101-SB01	SAM	03/10/25 10:58		Mohan	OK
17	Q1488-04	ENV-101-SB02	SAM	03/10/25 11:00		Mohan	OK
18	Q1488-06	ENV-102-SB01	SAM	03/10/25 11:02		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134958

Review By	Mohan	Review On	3/11/2025 9:32:51 AM
Supervise By	jaswal	Supervise On	3/11/2025 9:38:08 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84793,MP84794,MP84795,MP84796,MP84797,MP84798		
ICV Standard	MP84799		
CCV Standard	MP84801		
ICSA Standard			
CRI Standard	MP84803		
LCS Standard			
Chk Standard	MP84800,MP84802,MP84804,MP84826		

19	Q1488-08	ENV-102-SB02	SAM	03/10/25 11:05		Mohan	OK
20	Q1488-10	ENV-104-SB01	SAM	03/10/25 11:07		Mohan	OK
21	CCV76	CCV76	CCV	03/10/25 11:09		Mohan	OK
22	CCB76	CCB76	CCB	03/10/25 11:11		Mohan	OK
23	Q1488-12	ENV-104-SB02	SAM	03/10/25 11:14		Mohan	OK
24	Q1490-02	CLAY-SLUDGE-DRUM	SAM	03/10/25 11:16		Mohan	OK
25	Q1494-04	ASPHALT-SOIL	SAM	03/10/25 11:18		Mohan	OK
26	Q1494-06	SOIL	SAM	03/10/25 11:20		Mohan	OK
27	Q1494-08	SOIL-COMP	SAM	03/10/25 11:23		Mohan	OK
28	Q1510-01	FMC-25-0001-0005	SAM	03/10/25 11:25		Mohan	OK
29	Q1514-02	ENV-105-SB01	SAM	03/10/25 11:27		Mohan	OK
30	Q1514-04	ENV-105-SB02	SAM	03/10/25 11:30		Mohan	OK
31	Q1514-06	ENV-103-SB01	SAM	03/10/25 11:39		Mohan	OK
32	Q1514-06DUP	ENV-103-SB01DUP	DUP	03/10/25 11:41		Mohan	OK
33	CCV77	CCV77	CCV	03/10/25 11:48		Mohan	OK
34	CCB77	CCB77	CCB	03/10/25 11:53		Mohan	OK
35	Q1514-06MS	ENV-103-SB01MS	MS	03/10/25 11:55		Mohan	OK
36	Q1514-06MSD	ENV-103-SB01MSD	MSD	03/10/25 11:57		Mohan	OK
37	PB167034BL	PB167034BL	MB	03/10/25 11:59		Mohan	OK
38	PB167034BS	PB167034BS	LCS	03/10/25 12:02		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134958

Review By	Mohan	Review On	3/11/2025 9:32:51 AM
Supervise By	jaswal	Supervise On	3/11/2025 9:38:08 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84793,MP84794,MP84795,MP84796,MP84797,MP84798		
ICV Standard	MP84799		
CCV Standard	MP84801		
ICSA Standard			
CRI Standard	MP84803		
LCS Standard			
Chk Standard	MP84800,MP84802,MP84804,MP84826		

39	Q1478-01	IDW-AQ-MW-19B-CO	SAM	03/10/25 12:04		Mohan	OK
40	Q1478-01DUP	IDW-AQ-MW-19B-CO	DUP	03/10/25 12:06		Mohan	OK
41	Q1478-01MS	IDW-AQ-MW-19B-CO	MS	03/10/25 12:09		Mohan	OK
42	Q1478-01MSD	IDW-AQ-MW-19B-CO	MSD	03/10/25 12:11		Mohan	OK
43	Q1478-03	IDW-AQ-IW-01-COMP	SAM	03/10/25 12:13		Mohan	OK
44	Q1478-05	IDW-AQ-IW-02-COMP	SAM	03/10/25 12:15		Mohan	OK
45	CCV78	CCV78	CCV	03/10/25 12:18		Mohan	OK
46	CCB78	CCB78	CCB	03/10/25 12:26		Mohan	OK
47	Q1478-07	IDW-AQ-IW-03-COMP	SAM	03/10/25 12:28		Mohan	OK
48	Q1494-01	PURGE-WATER	SAM	03/10/25 12:30		Mohan	OK
49	Q1514-07	ENV-105-GW01	SAM	03/10/25 12:32	High	Mohan	Dilution
50	Q1514-08	ENV-103-GW01	SAM	03/10/25 12:35		Mohan	OK
51	Q1514-09	FB03062025	SAM	03/10/25 12:38		Mohan	OK
52	Q1516-03	GAS-AUD-25-0026	SAM	03/10/25 12:40		Mohan	OK
53	Q1517-01	MOO-25-0062	SAM	03/10/25 12:42		Mohan	OK
54	CCV79	CCV79	CCV	03/10/25 12:45		Mohan	OK
55	CCB79	CCB79	CCB	03/10/25 12:47		Mohan	OK
56	PB167020TB	PB167020TB	MB	03/10/25 12:49		Mohan	OK
57	Q1514-06L	ENV-103-SB01L	SD	03/10/25 12:55		Mohan	OK
58	Q1514-06A	ENV-103-SB01A	PS	03/10/25 12:57		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134958

Review By	Mohan	Review On	3/11/2025 9:32:51 AM
Supervise By	jaswal	Supervise On	3/11/2025 9:38:08 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84793,MP84794,MP84795,MP84796,MP84797,MP84798		
ICV Standard	MP84799		
CCV Standard	MP84801		
ICSA Standard			
CRI Standard	MP84803		
LCS Standard			
Chk Standard	MP84800,MP84802,MP84804,MP84826		

59	Q1478-01L	IDW-AQ-MW-19B-CO	SD	03/10/25 12:59		Mohan	OK
60	Q1478-01A	IDW-AQ-MW-19B-CO	PS	03/10/25 13:05		Mohan	OK
61	Q1514-07DL	ENV-105-GW01DL	SAM	03/10/25 13:08	Report 10X	Mohan	Confirms
62	CCV80	CCV80	CCV	03/10/25 13:10		Mohan	OK
63	CCB80	CCB80	CCB	03/10/25 13:12		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/12/25 10:27		Kareem	OK
2	S1	S1	CAL2	03/12/25 10:31		Kareem	OK
3	S2	S2	CAL3	03/12/25 10:35		Kareem	OK
4	S3	S3	CAL4	03/12/25 10:39		Kareem	OK
5	S4	S4	CAL5	03/12/25 10:44		Kareem	OK
6	S5	S5	CAL6	03/12/25 10:48		Kareem	OK
7	ICV01	ICV01	ICV	03/12/25 10:52	ICV As,Be,Tl,Zn (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	03/12/25 10:56		Kareem	OK
9	ICB01	ICB01	ICB	03/12/25 11:01		Kareem	OK
10	CRI01	CRI01	CRDL	03/12/25 11:05		Kareem	OK
11	ICSA01	ICSA01	ICSA	03/12/25 11:09		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	03/12/25 11:14		Kareem	OK
13	ICSADL	ICSADL	ICSA	03/12/25 11:21		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	03/12/25 11:25		Kareem	OK
15	CCV01	CCV01	CCV	03/12/25 11:29		Kareem	OK
16	CCB01	CCB01	CCB	03/12/25 11:34		Kareem	OK
17	PB167037BL	PB167037BL	MB	03/12/25 11:38		Kareem	OK
18	PB167037BS	PB167037BS	LCS	03/12/25 11:42	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84846
LCS Standard	
Chk Standard	MP84649,MP84650

19	PB166981BL	PB166981BL	MB	03/12/25 11:46		Kareem	OK
20	PB166981BS	PB166981BS	LCS	03/12/25 11:51	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
21	PB167038BL	PB167038BL	MB	03/12/25 11:55		Kareem	OK
22	PB167038BS	PB167038BS	LCS	03/12/25 11:59	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
23	PB167024BL	PB167024BL	MB	03/12/25 12:03		Kareem	OK
24	PB167024BS	PB167024BS	LCS	03/12/25 12:07	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
25	PB167100BL	PB167100BL	MB	03/12/25 12:11		Kareem	OK
26	PB167100BS	PB167100BS	LCS	03/12/25 12:16	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
27	CCV02	CCV02	CCV	03/12/25 12:20		Kareem	OK
28	CCB02	CCB02	CCB	03/12/25 12:24		Kareem	OK
29	PB166977BL	PB166977BL	MB	03/12/25 12:28		Kareem	OK
30	PB166977BS	PB166977BS	LCS	03/12/25 12:33	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
31	PB166963BL	PB166963BL	MB	03/12/25 12:37	Ba fail LCS	Kareem	Not Ok
32	PB166963BS	PB166963BS	LCS	03/12/25 12:41	Ba fail LCS	Kareem	Not Ok
33	Q1466-05	CITY WATER	SAM	03/12/25 12:49		Kareem	OK
34	Q1466-05DUP	CITY WATERDUP	DUP	03/12/25 12:53		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84846
LCS Standard	
Chk Standard	MP84649,MP84650

35	Q1466-05L	CITY WATERL	SD	03/12/25 12:58		Kareem	OK
36	Q1466-05MS	CITY WATERMS	MS	03/12/25 13:02	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
37	Q1466-05MSD	CITY WATERMSD	MSD	03/12/25 13:06	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
38	CCV03	CCV03	CCV	03/12/25 13:10		Kareem	OK
39	CCB03	CCB03	CCB	03/12/25 13:14		Kareem	OK
40	Q1466-05A	CITY WATERA	PS	03/12/25 13:18	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
41	Q1507-01	50-MIDDLESEX-AVE	SAM	03/12/25 13:22		Kareem	OK
42	Q1508-01	RBR251372	SAM	03/12/25 13:27		Kareem	OK
43	Q1510-01	FMC-25-0001-0005	SAM	03/12/25 13:31		Kareem	OK
44	PB167098BL	PB167098BL	MB	03/12/25 13:59		Kareem	OK
45	PB167098BS	PB167098BS	LCS	03/12/25 14:03	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
46	PB167099BL	PB167099BL	MB	03/12/25 14:07		Kareem	OK
47	PB167099BS	PB167099BS	LCS	03/12/25 14:12	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
48	Q1515-01	AU-06-030625	SAM	03/12/25 14:16		Kareem	OK
49	Q1515-01DUP	AU-06-030625DUP	DUP	03/12/25 14:20		Kareem	OK
50	CCV04	CCV04	CCV	03/12/25 14:24		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84846
LCS Standard	
Chk Standard	MP84649,MP84650

51	CCB04	CCB04	CCB	03/12/25 14:28		Kareem	OK
52	Q1540-01	OU4-CF-15R-031025	SAM	03/12/25 14:32		Kareem	OK
53	Q1540-01DUP	OU4-CF-15R-031025	DUP	03/12/25 14:37		Kareem	OK
54	Q1540-01L	OU4-CF-15R-031025	SD	03/12/25 14:41		Kareem	OK
55	Q1540-01A	OU4-CF-15R-031025	PS	03/12/25 14:58	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
56	PB167088TB	PB167088TB	MB	03/12/25 15:02		Kareem	OK
57	Q1545-01	BU-02-031125	SAM	03/12/25 15:06		Kareem	OK
58	Q1545-01DUP	BU-02-031125DUP	DUP	03/12/25 15:11		Kareem	OK
59	Q1545-01L	BU-02-031125L	SD	03/12/25 15:15		Kareem	OK
60	CCV05	CCV05	CCV	03/12/25 15:25		Kareem	OK
61	CCB05	CCB05	CCB	03/12/25 15:29		Kareem	OK
62	Q1545-01MS	BU-02-031125MS	MS	03/12/25 15:33	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
63	Q1545-01MSD	BU-02-031125MSD	MSD	03/12/25 15:37	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
64	Q1545-01A	BU-02-031125A	PS	03/12/25 15:41	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
65	Q1547-01	OR-620-JB-COMP-01	SAM	03/12/25 15:45		Kareem	OK
66	Q1547-06	OR-620-JB-COMP-02	SAM	03/12/25 15:49		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

67	Q1549-01	72-11978	SAM	03/12/25 15:54		Kareem	OK
68	Q1515-01L	AU-06-030625L	SD	03/12/25 15:58		Kareem	OK
69	Q1515-01MS	AU-06-030625MS	MS	03/12/25 16:02	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
70	Q1515-01MSD	AU-06-030625MSD	MSD	03/12/25 16:06	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
71	Q1515-01A	AU-06-030625A	PS	03/12/25 16:10	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
72	CCV06	CCV06	CCV	03/12/25 16:25		Kareem	OK
73	CCB06	CCB06	CCB	03/12/25 16:30		Kareem	OK
74	Q1540-01MS	OU4-CF-15R-031025	MS	03/12/25 16:34	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
75	Q1540-01MSD	OU4-CF-15R-031025	MSD	03/12/25 16:38	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
76	Q1516-03	GAS-AUD-25-0026	SAM	03/12/25 16:42		Kareem	OK
77	Q1519-01	WATER TREATMENT	SAM	03/12/25 16:47		Kareem	OK
78	LR1	LR1	HIGH STD	03/12/25 16:53		Kareem	OK
79	LR2	LR2	HIGH STD	03/12/25 16:58		Kareem	OK
80	Q1517-01DL	MOO-25-0062DL	SAM	03/12/25 17:02		Kareem	OK
81	Q1522-02	TW-WTS-04	SAM	03/12/25 17:07		Kareem	OK
82	Q1522-02DUP	TW-WTS-04DUP	DUP	03/12/25 17:11		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84846
LCS Standard	
Chk Standard	MP84649,MP84650

83	Q1522-02L	TW-WTS-04L	SD	03/12/25 17:16		Kareem	OK
84	CCV07	CCV07	CCV	03/12/25 17:25		Kareem	OK
85	CCB07	CCB07	CCB	03/12/25 17:30		Kareem	OK
86	Q1522-02MS	TW-WTS-04MS	MS	03/12/25 17:35	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
87	Q1522-02MSD	TW-WTS-04MSD	MSD	03/12/25 17:39	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
88	Q1522-02A	TW-WTS-04A	PS	03/12/25 17:43	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
89	Q1478-07	IDW-AQ-IW-03-COMF	SAM	03/12/25 17:47		Kareem	OK
90	Q1478-07DUP	IDW-AQ-IW-03-COMF	DUP	03/12/25 17:52		Kareem	OK
91	Q1478-07L	IDW-AQ-IW-03-COMF	SD	03/12/25 17:56		Kareem	OK
92	Q1478-07MS	IDW-AQ-IW-03-COMF	MS	03/12/25 18:00	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
93	Q1478-07MSD	IDW-AQ-IW-03-COMF	MSD	03/12/25 18:05	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
94	Q1478-07A	IDW-AQ-IW-03-COMF	PS	03/12/25 18:09	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
95	Q1478-01	IDW-AQ-MW-19B-CO	SAM	03/12/25 18:13		Kareem	OK
96	CCV08	CCV08	CCV	03/12/25 18:23		Kareem	OK
97	CCB08	CCB08	CCB	03/12/25 18:27		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135011

Review By	kareem	Review On	3/13/2025 10:36:08 AM
Supervise By	jaswal	Supervise On	3/14/2025 1:32:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

98	Q1478-03	IDW-AQ-IW-01-COMF	SAM	03/12/25 18:31		Kareem	OK
99	Q1478-05	IDW-AQ-IW-02-COMF	SAM	03/12/25 18:36		Kareem	OK
100	Q1477-01	RW7-SP100-2025022	SAM	03/12/25 18:40		Kareem	OK
101	Q1477-04	RW7-SP303-2025022	SAM	03/12/25 18:44		Kareem	OK
102	CCV09	CCV09	CCV	03/12/25 18:49		Kareem	OK
103	CCB09	CCB09	CCB	03/12/25 18:53		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135035

Review By	jaswal	Review On	3/14/2025 1:24:23 PM
Supervise By	Mohan	Supervise On	3/14/2025 1:33:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84846,MP84640,MP84639,MP84638,MP84637		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84640		
LCS Standard			
Chk Standard	MP84649,MP84650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/13/25 11:42		Kareem	OK
2	S1	S1	CAL2	03/13/25 11:46		Kareem	OK
3	S2	S2	CAL3	03/13/25 11:50		Kareem	OK
4	S3	S3	CAL4	03/13/25 11:54		Kareem	OK
5	S4	S4	CAL5	03/13/25 11:59		Kareem	OK
6	S5	S5	CAL6	03/13/25 12:03		Kareem	OK
7	ICV01	ICV01	ICV	03/13/25 12:07		Kareem	OK
8	LLICV01	LLICV01	LLICV	03/13/25 12:13		Kareem	OK
9	ICB01	ICB01	ICB	03/13/25 12:17		Kareem	OK
10	CRI01	CRI01	CRDL	03/13/25 12:21		Kareem	OK
11	ICSA01	ICSA01	ICSA	03/13/25 12:25		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	03/13/25 12:30		Kareem	OK
13	ICSADL	ICSADL	ICSA	03/13/25 12:34		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	03/13/25 12:38		Kareem	OK
15	CCV01	CCV01	CCV	03/13/25 12:53		Kareem	OK
16	CCB01	CCB01	CCB	03/13/25 12:58		Kareem	OK
17	PB166963BL	PB166963BL	MB	03/13/25 13:02		Kareem	OK
18	PB166963BS	PB166963BS	LCS	03/13/25 13:06		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135035

Review By	jaswal	Review On	3/14/2025 1:24:23 PM
Supervise By	Mohan	Supervise On	3/14/2025 1:33:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84846,MP84640,MP84639,MP84638,MP84637		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84640		
LCS Standard			
Chk Standard	MP84649,MP84650		

19	PB167116BL	PB167116BL	MB	03/13/25 13:10		Kareem	OK
20	PB167116BS	PB167116BS	LCS	03/13/25 13:15		Kareem	OK
21	Q1485-01	DN-B-41	SAM	03/13/25 13:19		Kareem	OK
22	Q1486-01	DN-B-40	SAM	03/13/25 13:23		Kareem	OK
23	Q1487-01	DN-B-42	SAM	03/13/25 13:27		Kareem	OK
24	Q1488-04	ENV-101-SB02	SAM	03/13/25 13:31		Kareem	OK
25	Q1488-06	ENV-102-SB01	SAM	03/13/25 13:35		Kareem	OK
26	Q1488-08	ENV-102-SB02	SAM	03/13/25 13:40		Kareem	OK
27	CCV02	CCV02	CCV	03/13/25 13:49		Kareem	OK
28	CCB02	CCB02	CCB	03/13/25 13:54		Kareem	OK
29	Q1488-10	ENV-104-SB01	SAM	03/13/25 13:58		Kareem	OK
30	Q1488-12	ENV-104-SB02	SAM	03/13/25 14:02		Kareem	OK
31	Q1488-14	ENV-104-GW01	SAM	03/13/25 14:07		Kareem	OK
32	Q1488-01	ENV-101-SB01	SAM	03/13/25 14:11		Kareem	OK
33	Q1488-03	ENV-101-SB02	SAM	03/13/25 14:15		Kareem	OK
34	Q1488-05	ENV-102-SB01	SAM	03/13/25 14:19		Kareem	OK
35	Q1488-07	ENV-102-SB02	SAM	03/13/25 14:23		Kareem	OK
36	Q1488-09	ENV-104-SB01	SAM	03/13/25 14:27		Kareem	OK
37	Q1488-11	ENV-104-SB02	SAM	03/13/25 14:31		Kareem	OK
38	Q1491-02	COMP	SAM	03/13/25 14:35		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135035

Review By	jaswal	Review On	3/14/2025 1:24:23 PM
Supervise By	Mohan	Supervise On	3/14/2025 1:33:41 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84846,MP84640,MP84639,MP84638,MP84637
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84640
LCS Standard	
Chk Standard	MP84649,MP84650

39	CCV03	CCV03	CCV	03/13/25 14:40		Kareem	OK
40	CCB03	CCB03	CCB	03/13/25 14:48		Kareem	OK
41	Q1495-01	001-WILLETS-PT-BL	SAM	03/13/25 14:53		Kareem	OK
42	Q1495-02	002-35TH-AVE(MAR)	SAM	03/13/25 14:57		Kareem	OK
43	Q1514-02	ENV-105-SB01	SAM	03/13/25 15:01		Kareem	OK
44	Q1514-04	ENV-105-SB02	SAM	03/13/25 15:06		Kareem	OK
45	Q1514-06	ENV-103-SB01	SAM	03/13/25 15:10		Kareem	OK
46	Q1514-06DUP	ENV-103-SB01DUP	DUP	03/13/25 15:15		Kareem	OK
47	Q1514-06L	ENV-103-SB01L	SD	03/13/25 15:19		Kareem	OK
48	Q1514-06MS	ENV-103-SB01MS	MS	03/13/25 15:23		Kareem	OK
49	Q1514-06MSD	ENV-103-SB01MSD	MSD	03/13/25 15:28		Kareem	OK
50	Q1514-06A	ENV-103-SB01A	PS	03/13/25 15:34		Kareem	OK
51	CCV04	CCV04	CCV	03/13/25 15:38		Kareem	OK
52	CCB04	CCB04	CCB	03/13/25 15:42		Kareem	OK
53	Q1514-08	ENV-103-GW01	SAM	03/13/25 15:46		Kareem	OK
54	Q1514-09	FB03062025	SAM	03/13/25 15:51		Kareem	OK
55	Q1514-01	ENV-105-SB01	SAM	03/13/25 15:55		Kareem	OK
56	Q1514-03	ENV-105-SB02	SAM	03/13/25 15:59		Kareem	OK
57	Q1514-05	ENV-103-SB01	SAM	03/13/25 16:03		Kareem	OK
58	Q1525-01	MW10	SAM	03/13/25 16:08	Mn high	Kareem	Dilution

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135035

Review By	jaswal	Review On	3/14/2025 1:24:23 PM
Supervise By	Mohan	Supervise On	3/14/2025 1:33:41 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84846,MP84640,MP84639,MP84638,MP84637
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84640
LCS Standard	
Chk Standard	MP84649,MP84650

59	Q1552-01	NB-418-JB-COMP-01	SAM	03/13/25 16:12		Kareem	OK
60	Q1553-01	RT2321	SAM	03/13/25 16:16		Kareem	OK
61	Q1556-01	OR-02-031225	SAM	03/13/25 16:20	MS,MSD fail for many parameters	Kareem	Not Ok
62	Q1556-01DUP	OR-02-031225DUP	DUP	03/13/25 16:24	MS,MSD fail for many parameters	Kareem	Not Ok
63	CCV05	CCV05	CCV	03/13/25 16:28		Kareem	OK
64	CCB05	CCB05	CCB	03/13/25 16:32		Kareem	OK
65	Q1556-01L	OR-02-031225L	SD	03/13/25 16:37	MS,MSD fail for many parameters	Kareem	Not Ok
66	Q1556-01MS	OR-02-031225MS	MS	03/13/25 16:41	Fail for many parameters	Kareem	Not Ok
67	Q1556-01MSD	OR-02-031225MSD	MSD	03/13/25 16:45	Fail for many parameters	Kareem	Not Ok
68	Q1556-01A	OR-02-031225A	PS	03/13/25 16:49	MS,MSD fail for many parameters	Kareem	Not Ok
69	LR1	LR1	HIGH STD	03/13/25 16:53		Kareem	OK
70	LR2	LR2	HIGH STD	03/13/25 16:58		Kareem	OK
71	PB167103TB	PB167103TB	MB	03/13/25 17:04		Kareem	OK
72	Q1547-05	OR-620-JB-COMP-01	SAM	03/13/25 17:08		Kareem	OK
73	Q1547-05DUP	OR-620-JB-COMP-01	DUP	03/13/25 17:13		Kareem	OK
74	Q1547-05L	OR-620-JB-COMP-01	SD	03/13/25 17:17		Kareem	OK
75	CCV06	CCV06	CCV	03/13/25 17:21		Kareem	OK
76	CCB06	CCB06	CCB	03/13/25 17:25		Kareem	OK
77	Q1547-05MS	OR-620-JB-COMP-01	MS	03/13/25 17:30		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135035

Review By	jaswal	Review On	3/14/2025 1:24:23 PM
Supervise By	Mohan	Supervise On	3/14/2025 1:33:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84846,MP84640,MP84639,MP84638,MP84637		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84640		
LCS Standard			
Chk Standard	MP84649,MP84650		

78	Q1547-05MSD	OR-620-JB-COMP-01	MSD	03/13/25 17:34		Kareem	OK
79	Q1547-05A	OR-620-JB-COMP-01	PS	03/13/25 17:38		Kareem	OK
80	Q1547-10	OR-620-JB-COMP-02	SAM	03/13/25 17:42		Kareem	OK
81	Q1552-05	NB-418-JB-COMP-01	SAM	03/13/25 17:47		Kareem	OK
82	PB167117BL	PB167117BL	MB	03/13/25 17:51		Kareem	OK
83	PB167117BS	PB167117BS	LCS	03/13/25 17:55		Kareem	OK
84	Q1525-01DL	MW10DL	SAM	03/13/25 17:59	5X for Mn	Kareem	Confirms
85	Q1488-13DL	ENV-102-GW01DL	SAM	03/13/25 18:04	10X for Ag	Kareem	Confirms
86	Q1514-07DL	ENV-105-GW01DL	SAM	03/13/25 18:08	10X for Ca,Ag,Fe	Kareem	Confirms
87	CCV07	CCV07	CCV	03/13/25 18:12		Kareem	OK
88	CCB07	CCB07	CCB	03/13/25 18:35		Kareem	OK
89	Q1488-13	ENV-102-GW01	SAM	03/13/25 18:39	Ag oversaturated	Kareem	Dilution
90	Q1514-07	ENV-105-GW01	SAM	03/13/25 18:44	Ag oversaturated,Ca,Fe high	Kareem	Dilution
91	Q1488-13REDL	ENV-102-GW01RE	SAM	03/13/25 18:49	Not used	Kareem	Not Ok
92	CCV08	CCV08	CCV	03/13/25 18:53		Kareem	OK
93	CCB08	CCB08	CCB	03/13/25 19:03		Kareem	OK
94	Q1528-02	22725	SAM	03/13/25 19:07		Kareem	OK
95	CCV09	CCV09	CCV	03/13/25 19:12		Kareem	OK
96	CCB09	CCB09	CCB	03/13/25 19:21		Kareem	OK