

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV35	Mercury	4.01	4.0	100	90 - 110	CV	03/10/2025	10:22	LB134958

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478

Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV75	Mercury	4.61	5.0	92	90 - 110	CV	03/10/2025	10:34	LB134958
CCV76	Mercury	4.87	5.0	97	90 - 110	CV	03/10/2025	11:09	LB134958
CCV77	Mercury	5.16	5.0	103	90 - 110	CV	03/10/2025	11:48	LB134958
CCV78	Mercury	5.20	5.0	104	90 - 110	CV	03/10/2025	12:18	LB134958
CCV79	Mercury	5.06	5.0	101	90 - 110	CV	03/10/2025	12:45	LB134958
CCV80	Mercury	5.43	5.0	109	90 - 110	CV	03/10/2025	13:10	LB134958

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1050	1000	105	95 - 105	P	03/12/2025	10:52	LB135011
	Barium	537	520	103	95 - 105	P	03/12/2025	10:52	LB135011
	Cadmium	506	510	99	95 - 105	P	03/12/2025	10:52	LB135011
	Chromium	541	520	104	95 - 105	P	03/12/2025	10:52	LB135011
	Lead	1020	1000	102	95 - 105	P	03/12/2025	10:52	LB135011
	Selenium	1050	1000	105	95 - 105	P	03/12/2025	10:52	LB135011
	Silver	262	250	105	95 - 105	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. SDG No.: Q1478

Contract: JACO05 Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	22.2	20.0	111	80 - 120	P	03/12/2025	10:56	LB135011
	Barium	89.2	100	89	80 - 120	P	03/12/2025	10:56	LB135011
	Cadmium	5.83	6.0	97	80 - 120	P	03/12/2025	10:56	LB135011
	Chromium	10.6	10.0	106	80 - 120	P	03/12/2025	10:56	LB135011
	Lead	11.7	12.0	97	80 - 120	P	03/12/2025	10:56	LB135011
	Selenium	17.5	20.0	87	80 - 120	P	03/12/2025	10:56	LB135011
	Silver	11.0	10.0	110	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478

Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5210	5000	104	90 - 110	P	03/12/2025	11:29	LB135011
	Barium	9300	10000	93	90 - 110	P	03/12/2025	11:29	LB135011
	Cadmium	2470	2500	99	90 - 110	P	03/12/2025	11:29	LB135011
	Chromium	1050	1000	105	90 - 110	P	03/12/2025	11:29	LB135011
	Lead	5060	5000	101	90 - 110	P	03/12/2025	11:29	LB135011
	Selenium	5000	5000	100	90 - 110	P	03/12/2025	11:29	LB135011
	Silver	1260	1250	101	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Arsenic	5150	5000	103	90 - 110	P	03/12/2025	12:20	LB135011
	Barium	9180	10000	92	90 - 110	P	03/12/2025	12:20	LB135011
	Cadmium	2480	2500	99	90 - 110	P	03/12/2025	12:20	LB135011
	Chromium	1070	1000	107	90 - 110	P	03/12/2025	12:20	LB135011
	Lead	5010	5000	100	90 - 110	P	03/12/2025	12:20	LB135011
	Selenium	5180	5000	104	90 - 110	P	03/12/2025	12:20	LB135011
	Silver	1280	1250	103	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Arsenic	5100	5000	102	90 - 110	P	03/12/2025	13:10	LB135011
	Barium	9270	10000	93	90 - 110	P	03/12/2025	13:10	LB135011
	Cadmium	2480	2500	99	90 - 110	P	03/12/2025	13:10	LB135011
	Chromium	1060	1000	106	90 - 110	P	03/12/2025	13:10	LB135011
	Lead	5000	5000	100	90 - 110	P	03/12/2025	13:10	LB135011
	Selenium	5090	5000	102	90 - 110	P	03/12/2025	13:10	LB135011
	Silver	1290	1250	103	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Arsenic	5330	5000	107	90 - 110	P	03/12/2025	14:24	LB135011
	Barium	9270	10000	93	90 - 110	P	03/12/2025	14:24	LB135011
	Cadmium	2500	2500	100	90 - 110	P	03/12/2025	14:24	LB135011
	Chromium	1060	1000	106	90 - 110	P	03/12/2025	14:24	LB135011
	Lead	5090	5000	102	90 - 110	P	03/12/2025	14:24	LB135011
	Selenium	5140	5000	103	90 - 110	P	03/12/2025	14:24	LB135011
	Silver	1280	1250	103	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Arsenic	5210	5000	104	90 - 110	P	03/12/2025	15:25	LB135011
	Barium	9220	10000	92	90 - 110	P	03/12/2025	15:25	LB135011
	Cadmium	2460	2500	98	90 - 110	P	03/12/2025	15:25	LB135011
	Chromium	1050	1000	105	90 - 110	P	03/12/2025	15:25	LB135011
	Lead	5010	5000	100	90 - 110	P	03/12/2025	15:25	LB135011
	Selenium	5000	5000	100	90 - 110	P	03/12/2025	15:25	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478

Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1250	1250	100	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Arsenic	5190	5000	104	90 - 110	P	03/12/2025	16:25	LB135011
	Barium	9100	10000	91	90 - 110	P	03/12/2025	16:25	LB135011
	Cadmium	2450	2500	98	90 - 110	P	03/12/2025	16:25	LB135011
	Chromium	1050	1000	105	90 - 110	P	03/12/2025	16:25	LB135011
	Lead	4990	5000	100	90 - 110	P	03/12/2025	16:25	LB135011
	Selenium	4960	5000	99	90 - 110	P	03/12/2025	16:25	LB135011
	Silver	1260	1250	100	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Arsenic	5300	5000	106	90 - 110	P	03/12/2025	17:25	LB135011
	Barium	9240	10000	92	90 - 110	P	03/12/2025	17:25	LB135011
	Cadmium	2460	2500	98	90 - 110	P	03/12/2025	17:25	LB135011
	Chromium	1050	1000	105	90 - 110	P	03/12/2025	17:25	LB135011
	Lead	5030	5000	101	90 - 110	P	03/12/2025	17:25	LB135011
	Selenium	5120	5000	102	90 - 110	P	03/12/2025	17:25	LB135011
	Silver	1270	1250	102	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Arsenic	5240	5000	105	90 - 110	P	03/12/2025	18:23	LB135011
	Barium	9130	10000	91	90 - 110	P	03/12/2025	18:23	LB135011
	Cadmium	2490	2500	99	90 - 110	P	03/12/2025	18:23	LB135011
	Chromium	1070	1000	107	90 - 110	P	03/12/2025	18:23	LB135011
	Lead	5060	5000	101	90 - 110	P	03/12/2025	18:23	LB135011
	Selenium	5010	5000	100	90 - 110	P	03/12/2025	18:23	LB135011
	Silver	1280	1250	102	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Arsenic	5290	5000	106	90 - 110	P	03/12/2025	18:49	LB135011
	Barium	9220	10000	92	90 - 110	P	03/12/2025	18:49	LB135011
	Cadmium	2500	2500	100	90 - 110	P	03/12/2025	18:49	LB135011
	Chromium	1050	1000	105	90 - 110	P	03/12/2025	18:49	LB135011
	Lead	5060	5000	101	90 - 110	P	03/12/2025	18:49	LB135011
	Selenium	5060	5000	101	90 - 110	P	03/12/2025	18:49	LB135011
	Silver	1270	1250	102	90 - 110	P	03/12/2025	18:49	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. SDG No.: Q1478
 Contract: JACO05 Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1020	1000	102	90 - 110	P	03/13/2025	12:07	LB135035
	Barium	533	520	102	90 - 110	P	03/13/2025	12:07	LB135035
	Cadmium	503	510	99	90 - 110	P	03/13/2025	12:07	LB135035
	Chromium	525	520	101	90 - 110	P	03/13/2025	12:07	LB135035
	Lead	1000	1000	100	90 - 110	P	03/13/2025	12:07	LB135035
	Selenium	1040	1000	104	90 - 110	P	03/13/2025	12:07	LB135035
	Silver	252	250	101	90 - 110	P	03/13/2025	12:07	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. SDG No.: Q1478

Contract: JACO05 Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	17.3	20.0	86	80 - 120	P	03/13/2025	12:13	LB135035
	Barium	92.2	100	92	80 - 120	P	03/13/2025	12:13	LB135035
	Cadmium	5.95	6.0	99	80 - 120	P	03/13/2025	12:13	LB135035
	Chromium	10.1	10.0	101	80 - 120	P	03/13/2025	12:13	LB135035
	Lead	12.6	12.0	105	80 - 120	P	03/13/2025	12:13	LB135035
	Selenium	19.2	20.0	96	80 - 120	P	03/13/2025	12:13	LB135035
	Silver	10.6	10.0	106	80 - 120	P	03/13/2025	12:13	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478

Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5160	5000	103	90 - 110	P	03/13/2025	12:53	LB135035
	Barium	9860	10000	99	90 - 110	P	03/13/2025	12:53	LB135035
	Cadmium	2570	2500	103	90 - 110	P	03/13/2025	12:53	LB135035
	Chromium	1050	1000	105	90 - 110	P	03/13/2025	12:53	LB135035
	Lead	5190	5000	104	90 - 110	P	03/13/2025	12:53	LB135035
	Selenium	5070	5000	101	90 - 110	P	03/13/2025	12:53	LB135035
	Silver	1300	1250	104	90 - 110	P	03/13/2025	12:53	LB135035
CCV02	Arsenic	5290	5000	106	90 - 110	P	03/13/2025	13:49	LB135035
	Barium	10100	10000	101	90 - 110	P	03/13/2025	13:49	LB135035
	Cadmium	2630	2500	105	90 - 110	P	03/13/2025	13:49	LB135035
	Chromium	1100	1000	110	90 - 110	P	03/13/2025	13:49	LB135035
	Lead	5330	5000	107	90 - 110	P	03/13/2025	13:49	LB135035
	Selenium	5180	5000	104	90 - 110	P	03/13/2025	13:49	LB135035
	Silver	1310	1250	105	90 - 110	P	03/13/2025	13:49	LB135035
CCV03	Arsenic	5130	5000	102	90 - 110	P	03/13/2025	14:40	LB135035
	Barium	9830	10000	98	90 - 110	P	03/13/2025	14:40	LB135035
	Cadmium	2540	2500	102	90 - 110	P	03/13/2025	14:40	LB135035
	Chromium	1050	1000	106	90 - 110	P	03/13/2025	14:40	LB135035
	Lead	5150	5000	103	90 - 110	P	03/13/2025	14:40	LB135035
	Selenium	5060	5000	101	90 - 110	P	03/13/2025	14:40	LB135035
	Silver	1280	1250	103	90 - 110	P	03/13/2025	14:40	LB135035
CCV04	Arsenic	4950	5000	99	90 - 110	P	03/13/2025	15:38	LB135035
	Barium	9580	10000	96	90 - 110	P	03/13/2025	15:38	LB135035
	Cadmium	2510	2500	100	90 - 110	P	03/13/2025	15:38	LB135035
	Chromium	1030	1000	103	90 - 110	P	03/13/2025	15:38	LB135035
	Lead	5050	5000	101	90 - 110	P	03/13/2025	15:38	LB135035
	Selenium	4830	5000	97	90 - 110	P	03/13/2025	15:38	LB135035
	Silver	1240	1250	99	90 - 110	P	03/13/2025	15:38	LB135035
CCV05	Arsenic	5250	5000	105	90 - 110	P	03/13/2025	16:28	LB135035
	Barium	9880	10000	99	90 - 110	P	03/13/2025	16:28	LB135035
	Cadmium	2630	2500	105	90 - 110	P	03/13/2025	16:28	LB135035
	Chromium	1070	1000	107	90 - 110	P	03/13/2025	16:28	LB135035
	Lead	5300	5000	106	90 - 110	P	03/13/2025	16:28	LB135035
	Selenium	5150	5000	103	90 - 110	P	03/13/2025	16:28	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478

Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1300	1250	104	90 - 110	P	03/13/2025	16:28	LB135035
CCV06	Arsenic	5070	5000	101	90 - 110	P	03/13/2025	17:21	LB135035
	Barium	9980	10000	100	90 - 110	P	03/13/2025	17:21	LB135035
	Cadmium	2570	2500	103	90 - 110	P	03/13/2025	17:21	LB135035
	Chromium	1060	1000	106	90 - 110	P	03/13/2025	17:21	LB135035
	Lead	5180	5000	104	90 - 110	P	03/13/2025	17:21	LB135035
	Selenium	4970	5000	100	90 - 110	P	03/13/2025	17:21	LB135035
	Silver	1290	1250	103	90 - 110	P	03/13/2025	17:21	LB135035
CCV07	Arsenic	5250	5000	105	90 - 110	P	03/13/2025	18:12	LB135035
	Barium	9960	10000	100	90 - 110	P	03/13/2025	18:12	LB135035
	Cadmium	2640	2500	106	90 - 110	P	03/13/2025	18:12	LB135035
	Chromium	1090	1000	109	90 - 110	P	03/13/2025	18:12	LB135035
	Lead	5330	5000	107	90 - 110	P	03/13/2025	18:12	LB135035
	Selenium	5180	5000	104	90 - 110	P	03/13/2025	18:12	LB135035
	Silver	1310	1250	105	90 - 110	P	03/13/2025	18:12	LB135035
CCV08	Arsenic	4950	5000	99	90 - 110	P	03/13/2025	18:53	LB135035
	Barium	9390	10000	94	90 - 110	P	03/13/2025	18:53	LB135035
	Cadmium	2510	2500	100	90 - 110	P	03/13/2025	18:53	LB135035
	Chromium	1040	1000	104	90 - 110	P	03/13/2025	18:53	LB135035
	Lead	5060	5000	101	90 - 110	P	03/13/2025	18:53	LB135035
	Selenium	4840	5000	97	90 - 110	P	03/13/2025	18:53	LB135035
	Silver	1250	1250	100	90 - 110	P	03/13/2025	18:53	LB135035
CCV09	Arsenic	5280	5000	106	90 - 110	P	03/13/2025	19:12	LB135035
	Barium	9620	10000	96	90 - 110	P	03/13/2025	19:12	LB135035
	Cadmium	2550	2500	102	90 - 110	P	03/13/2025	19:12	LB135035
	Chromium	1070	1000	107	90 - 110	P	03/13/2025	19:12	LB135035
	Lead	5120	5000	102	90 - 110	P	03/13/2025	19:12	LB135035
	Selenium	5440	5000	109	90 - 110	P	03/13/2025	19:12	LB135035
	Silver	1290	1250	104	90 - 110	P	03/13/2025	19:12	LB135035