

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Sample ID:** Q1478-01 **Client ID:** IDW-AQ-MW-19B-COMP-022825DUP
Percent Solids for Sample: NA **Duplicate ID** Q1478-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Sample ID:** Q1478-01MS **Client ID:** IDW-AQ-MW-19B-COMP-022825MSD
Percent Solids for Sample: NA **Duplicate ID** Q1478-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.54		3.47		2		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Sample ID:** Q1478-07 **Client ID:** IDW-AQ-IW-03-COMP-022825DUP
Percent Solids for Sample: NA **Duplicate ID** Q1478-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	112		116		4		P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	14.9		14.6		2		P
Lead	ug/L	20	6.00	U	6.00	U			P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Sample ID:** Q1478-07MS **Client ID:** IDW-AQ-IW-03-COMP-022825MSD
Percent Solids for Sample: NA **Duplicate ID** Q1478-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	415		422		2		P
Barium	ug/L	20	202		194		4		P
Cadmium	ug/L	20	94.5		96.0		2		P
Chromium	ug/L	20	225		229		2		P
Lead	ug/L	20	461		467		1		P
Selenium	ug/L	20	949		968		2		P
Silver	ug/L	20	13.3		13.1		2		P